

Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_B\Data\FB013123\  
 Data File : FB029190.D  
 Signal(s) : FID2B.CH  
 Acq On : 31 Jan 2023 10:03  
 Operator : AJ/MA  
 Sample : 10 GRO STD  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 FID\_B  
 ClientSampleId :  
 10 GRO STD

Integration File: Calibration.e  
 Quant Time: Jan 31 11:15:19 2023  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_B\Method\FB013123.M  
 Quant Title :  
 QLast Update : Tue Jan 31 11:15:04 2023  
 Response via : Initial Calibration  
 Integrator: ChemStation

Volume Inj. : 5 g/ml  
 Signal Phase : RTX-502.2  
 Signal Info : 60mx0.53mmx3.00um

| Compound                     | R.T.   | Response | Conc Units   |
|------------------------------|--------|----------|--------------|
| -----                        |        |          |              |
| System Monitoring Compounds  |        |          |              |
| 5) s AAA-TFT                 | 8.788  | 130421   | 11.049 ng/ml |
| Target Compounds             |        |          |              |
| 1) t 2-Methylpentane         | 4.707  | 177481   | 15.770 ng/ml |
| 2) t 2,2,4-Trimethylpentane  | 7.414  | 209770   | 15.184 ng/ml |
| 3) t n-Heptane               | 7.748  | 62107    | 5.143 ng/ml  |
| 4) t Benzene                 | 7.883  | 89561    | 5.216 ng/ml  |
| 6) t Toluene                 | 10.617 | 268168   | 16.021 ng/ml |
| 7) t Ethylbenzene            | 13.053 | 86811    | 5.245 ng/ml  |
| 8) t m-Xylene                | 13.186 | 171632   | 10.401 ng/ml |
| 9) t O-Xylene                | 13.912 | 171086   | 10.411 ng/ml |
| 10) t 1,2,4-Trimethylbenzene | 16.187 | 175039   | 9.973 ng/ml  |
| -----                        |        |          |              |

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_B\Data\FB013123\  
Data File : FB029190.D  
Signal(s) : FID2B.CH  
Acq On : 31 Jan 2023 10:03  
Operator : AJ/MA  
Sample : 10 GRO STD  
Misc :  
ALS Vial : 2 Sample Multiplier: 1

Instrument :  
FID\_B  
ClientSampleId :  
10 GRO STD

Integration File: Calibration.e  
Quant Time: Jan 31 11:15:19 2023  
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_B\Method\FB013123.M  
Quant Title :  
QLast Update : Tue Jan 31 11:15:04 2023  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 5 g/ml  
Signal Phase : RTX-502.2  
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane


R.T.: 4.707 min  
Delta R.T.: 0.000 min  
Response: 177481  
Conc: 15.77 ng/ml

Instrument :  
FID\_B  
ClientSampleId :  
10 GRO STD




#2 2,2,4-Trimethylpentane

R.T.: 7.414 min  
Delta R.T.: -0.003 min  
Response: 209770  
Conc: 15.18 ng/ml




#3 n-Heptane

R.T.: 7.748 min  
Delta R.T.: 0.001 min  
Response: 62107  
Conc: 5.14 ng/ml



#4 Benzene


R.T.: 7.883 min  
Delta R.T.: -0.002 min  
Response: 89561  
Conc: 5.22 ng/ml



#5 AAA-TFT


R.T.: 8.788 min  
Delta R.T.: 0.000 min  
Response: 130421  
Conc: 11.05 ng/ml

Instrument :  
FID\_B  
ClientSampleId :  
10 GRO STD




#6 Toluene

R.T.: 10.617 min  
Delta R.T.: 0.000 min  
Response: 268168  
Conc: 16.02 ng/ml




#7 Ethylbenzene

R.T.: 13.053 min  
Delta R.T.: 0.000 min  
Response: 86811  
Conc: 5.24 ng/ml



#8 m-Xylene


R.T.: 13.186 min  
Delta R.T.: 0.000 min  
Response: 171632  
Conc: 10.40 ng/ml



#9 O-Xylene

R.T.: 13.912 min  
Delta R.T.: 0.000 min  
Response: 171086  
Conc: 10.41 ng/ml

Instrument :  
FID\_B  
ClientSampleId :  
10 GRO STD



#10 1,2,4-Trimethylbenzene

R.T.: 16.187 min  
Delta R.T.: 0.000 min  
Response: 175039  
Conc: 9.97 ng/ml