

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB020119\
 Data File : FB019610.D
 Signal(s) : FID2B.CH
 Acq On : 1 Feb 2019 13:07
 Operator : MJ/SJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Feb 01 13:48:22 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB012919.M
 Quant Title :
 QLast Update : Tue Jan 29 12:07:12 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	9.041	188901	16.890 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.896	491971	26.850 ng/ml
2) t 2,2,4-Trimethylpentane	7.657	473785	27.123 ng/ml
3) t n-Heptane	7.997	148045	8.892 ng/ml
4) t Benzene	8.135	170964	8.879 ng/ml
6) t Toluene	10.890	495633	28.477 ng/ml
7) t Ethylbenzene	13.341	200852	9.276 ng/ml
8) t m-Xylene	13.475	418398	18.525 ng/ml
9) t O-Xylene	14.206	410494	18.563 ng/ml
10) t 1,2,4-Trimethylbenzene	16.493	418173	18.373 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB020119\
Data File : FB019610.D
Signal(s) : FID2B.CH
Acq On : 1 Feb 2019 13:07
Operator : MJ/SJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 8 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Feb 01 13:48:22 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB012919.M
Quant Title :
QLast Update : Tue Jan 29 12:07:12 2019
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.896 min
Delta R.T.: -0.006 min
Response: 491971
Conc: 26.85 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.657 min
Delta R.T.: -0.007 min
Response: 473785
Conc: 27.12 ng/ml




#3 n-Heptane

R.T.: 7.997 min
Delta R.T.: -0.007 min
Response: 148045
Conc: 8.89 ng/ml




#4 Benzene

R.T.: 8.135 min
Delta R.T.: -0.006 min
Response: 170964
Conc: 8.88 ng/ml




#5 AAA-TFT

R.T.: 9.041 min
Delta R.T.: -0.008 min
Response: 188901
Conc: 16.89 ng/ml




#6 Toluene

R.T.: 10.890 min
Delta R.T.: -0.007 min
Response: 495633
Conc: 28.48 ng/ml




#7 Ethylbenzene

R.T.: 13.341 min
Delta R.T.: -0.007 min
Response: 200852
Conc: 9.28 ng/ml




#8 m-Xylene

R.T.: 13.475 min
Delta R.T.: -0.007 min
Response: 418398
Conc: 18.52 ng/ml



#9 O-Xylene

R.T.: 14.206 min
Delta R.T.: -0.008 min
Response: 410494
Conc: 18.56 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.493 min
Delta R.T.: -0.007 min
Response: 418173
Conc: 18.37 ng/ml