

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB020420\
Data File : FB023615.D
Signal(s) : FID2B.CH
Acq On : 4 Feb 2020 11:28
Operator : DD/AJ
Sample : L1368-02
Misc : 5.08g/5.00ml DI WATER
ALS Vial : 6 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Feb 04 14:42:09 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB011620.M
Quant Title :
QLast Update : Thu Jan 16 11:42:55 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um

Compound		R.T.	Response	Conc Units

System Monitoring Compounds				
5) s	AAA-TFT	8.799	462304	19.462 ng/ml
Target Compounds				
1) t	2-Methylpentane	4.686	10131	0.321 ng/ml
6) t	Toluene	10.631	66208	1.675 ng/ml
7) t	Ethylbenzene	13.067	19798	0.530 ng/ml
8) t	m-Xylene	13.202	45244	1.272 ng/ml
9) t	O-Xylene	13.933	60209	1.742 ng/ml
10) t	1,2,4-Trimethylbenzene	16.210	321054	10.980 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB020420\
Data File : FB023615.D
Signal(s) : FID2B.CH
Acq On : 4 Feb 2020 11:28
Operator : DD/AJ
Sample : L1368-02
Misc : 5.08g/5.00ml DI WATER
ALS Vial : 6 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Feb 04 14:42:09 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB011620.M
Quant Title :
QLast Update : Thu Jan 16 11:42:55 2020
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.686 min
Delta R.T.: -0.001 min
Response: 10131
Conc: 0.32 ng/ml




#5 AAA-TFT

R.T.: 8.799 min
Delta R.T.: 0.003 min
Response: 462304
Conc: 19.46 ng/ml




#6 Toluene

R.T.: 10.631 min
Delta R.T.: 0.003 min
Response: 66208
Conc: 1.68 ng/ml




#7 Ethylbenzene

R.T.: 13.067 min
Delta R.T.: -0.001 min
Response: 19798
Conc: 0.53 ng/ml




#8 m-Xylene

R.T.: 13.202 min
Delta R.T.: 0.000 min
Response: 45244
Conc: 1.27 ng/ml



#9 O-Xylene

R.T.: 13.933 min
Delta R.T.: 0.004 min
Response: 60209
Conc: 1.74 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.210 min
Delta R.T.: 0.003 min
Response: 321054
Conc: 10.98 ng/ml