

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB020620\
 Data File : FB023632.D
 Signal(s) : FID2B.CH
 Acq On : 6 Feb 2020 10:07
 Operator : DD/AJ
 Sample : 50 GRO STD
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :
 50 GRO STD

Integration File: Calibration.e
 Quant Time: Feb 06 10:26:42 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB020620.M
 Quant Title :
 QLast Update : Thu Feb 06 10:19:14 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.798	1041830	49.598 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.686	2416482	78.504 ng/ml
2) t 2,2,4-Trimethylpentane	7.416	2697162	81.852 ng/ml
3) t n-Heptane	7.756	885156	27.169 ng/ml
4) t Benzene	7.891	1072796	25.923 ng/ml
6) t Toluene	10.630	3066664	75.598 ng/ml
7) t Ethylbenzene	13.070	958399	25.208 ng/ml
8) t m-Xylene	13.203	1812217	49.055 ng/ml
9) t O-Xylene	13.931	1726233	48.043 ng/ml
10) t 1,2,4-Trimethylbenzene	16.210	1524658	51.162 ng/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB020620\
Data File : FB023632.D
Signal(s) : FID2B.CH
Acq On : 6 Feb 2020 10:07
Operator : DD/AJ
Sample : 50 GRO STD
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :
50 GRO STD

Integration File: Calibration.e
Quant Time: Feb 06 10:26:42 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB020620.M
Quant Title :
QLast Update : Thu Feb 06 10:19:14 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane


R.T.: 4.686 min
Delta R.T.: -0.003 min
Response: 2416482
Conc: 78.50 ng/ml

Instrument :
FID_B
ClientSampleId :
50 GRO STD




#2 2,2,4-Trimethylpentane

R.T.: 7.416 min
Delta R.T.: -0.002 min
Response: 2697162
Conc: 81.85 ng/ml




#3 n-Heptane

R.T.: 7.756 min
Delta R.T.: -0.002 min
Response: 885156
Conc: 27.17 ng/ml



#4 Benzene


R.T.: 7.891 min
Delta R.T.: 0.000 min
Response: 1072796
Conc: 25.92 ng/ml



#5 AAA-TFT


R.T.: 8.798 min
Delta R.T.: 0.000 min
Response: 1041830
Conc: 49.60 ng/ml

Instrument :
FID_B
ClientSampleId :
50 GRO STD




#6 Toluene

R.T.: 10.630 min
Delta R.T.: 0.000 min
Response: 3066664
Conc: 75.60 ng/ml




#7 Ethylbenzene

R.T.: 13.070 min
Delta R.T.: 0.000 min
Response: 958399
Conc: 25.21 ng/ml



#8 m-Xylene


R.T.: 13.203 min
Delta R.T.: 0.000 min
Response: 1812217
Conc: 49.05 ng/ml



#9 O-Xylene

R.T.: 13.931 min
Delta R.T.: 0.000 min
Response: 1726233
Conc: 48.04 ng/ml

Instrument :
FID_B
ClientSampleId :
50 GRO STD



#10 1,2,4-Trimethylbenzene

R.T.: 16.210 min
Delta R.T.: 0.000 min
Response: 1524658
Conc: 51.16 ng/ml