

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB021119\
 Data File : FB019681.D
 Signal(s) : FID2B.CH
 Acq On : 11 Feb 2019 9:47
 Operator : MJ/SJ
 Sample : BSF0211S1
 Misc : 5.02g/5.00ml DI WATER
 ALS Vial : 3 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Feb 12 00:27:55 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB012919.M
 Quant Title :
 QLast Update : Tue Jan 29 12:07:12 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	9.038	187020	15.999 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.894	562510	32.491 ng/ml
2) t 2,2,4-Trimethylpentane	7.652	558952	33.705 ng/ml
3) t n-Heptane	7.994	171764	10.659 ng/ml
4) t Benzene	8.130	192985	10.232 ng/ml
6) t Toluene	10.885	532149	30.422 ng/ml
7) t Ethylbenzene	13.337	218679	9.965 ng/ml
8) t m-Xylene	13.470	445334	19.457 ng/ml
9) t O-Xylene	14.202	419739	18.793 ng/ml
10) t 1,2,4-Trimethylbenzene	16.489	428714	18.503 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB021119\
Data File : FB019681.D
Signal(s) : FID2B.CH
Acq On : 11 Feb 2019 9:47
Operator : MJ/SJ
Sample : BSF0211S1
Misc : 5.02g/5.00ml DI WATER
ALS Vial : 3 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Feb 12 00:27:55 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB012919.M
Quant Title :
QLast Update : Tue Jan 29 12:07:12 2019
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.894 min
Delta R.T.: -0.008 min
Response: 562510
Conc: 32.49 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.652 min
Delta R.T.: -0.012 min
Response: 558952
Conc: 33.71 ng/ml




#3 n-Heptane

R.T.: 7.994 min
Delta R.T.: -0.011 min
Response: 171764
Conc: 10.66 ng/ml




#4 Benzene

R.T.: 8.130 min
Delta R.T.: -0.011 min
Response: 192985
Conc: 10.23 ng/ml




#5 AAA-TFT

R.T.: 9.038 min
Delta R.T.: -0.011 min
Response: 187020
Conc: 16.00 ng/ml




#6 Toluene

R.T.: 10.885 min
Delta R.T.: -0.012 min
Response: 532149
Conc: 30.42 ng/ml




#7 Ethylbenzene

R.T.: 13.337 min
Delta R.T.: -0.011 min
Response: 218679
Conc: 9.96 ng/ml




#8 m-Xylene

R.T.: 13.470 min
Delta R.T.: -0.012 min
Response: 445334
Conc: 19.46 ng/ml



#9 O-Xylene

R.T.: 14.202 min
Delta R.T.: -0.012 min
Response: 419739
Conc: 18.79 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.489 min
Delta R.T.: -0.012 min
Response: 428714
Conc: 18.50 ng/ml