

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB022020\
 Data File : FB023696.D
 Signal(s) : FID2B.CH
 Acq On : 20 Feb 2020 14:15
 Operator : DD/AJ
 Sample : L1606-01MSD
 Misc : 5.06g/5.00ml DI WATER
 ALS Vial : 6 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Feb 21 00:19:50 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB020620.M
 Quant Title :
 QLast Update : Thu Feb 06 11:13:47 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.794	456975	22.445 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.687	819631	27.239 ng/ml
2) t 2,2,4-Trimethylpentane	7.414	841712	25.795 ng/ml
3) t n-Heptane	7.753	272824	8.480 ng/ml
4) t Benzene	7.887	401709	9.822 ng/ml
6) t Toluene	10.625	1170826	29.496 ng/ml
7) t Ethylbenzene	13.065	359676	9.675 ng/ml
8) t m-Xylene	13.198	695239	19.468 ng/ml
9) t O-Xylene	13.926	662105	19.197 ng/ml
10) t 1,2,4-Trimethylbenzene	16.204	560858	19.156 ng/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB022020\
Data File : FB023696.D
Signal(s) : FID2B.CH
Acq On : 20 Feb 2020 14:15
Operator : DD/AJ
Sample : L1606-01MSD
Misc : 5.06g/5.00ml DI WATER
ALS Vial : 6 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Feb 21 00:19:50 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB020620.M
Quant Title :
QLast Update : Thu Feb 06 11:13:47 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um

