

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB022119\
 Data File : FB019869.D
 Signal(s) : FID2B.CH
 Acq On : 21 Feb 2019 20:48
 Operator : MJ/SJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Feb 22 00:56:34 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB012919.M
 Quant Title :
 QLast Update : Tue Jan 29 12:07:12 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	9.041	199530	17.069 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.897	596928	34.478 ng/ml
2) t 2,2,4-Trimethylpentane	7.656	572350	34.513 ng/ml
3) t n-Heptane	7.997	174920	10.855 ng/ml
4) t Benzene	8.134	194105	10.291 ng/ml
6) t Toluene	10.888	525635	30.049 ng/ml
7) t Ethylbenzene	13.340	208784	9.514 ng/ml
8) t m-Xylene	13.473	433135	18.924 ng/ml
9) t O-Xylene	14.205	433488	19.409 ng/ml
10) t 1,2,4-Trimethylbenzene	16.491	432946	18.686 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB022119\
Data File : FB019869.D
Signal(s) : FID2B.CH
Acq On : 21 Feb 2019 20:48
Operator : MJ/SJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 26 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Feb 22 00:56:34 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB012919.M
Quant Title :
QLast Update : Tue Jan 29 12:07:12 2019
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.897 min
Delta R.T.: -0.005 min
Response: 596928
Conc: 34.48 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.656 min
Delta R.T.: -0.009 min
Response: 572350
Conc: 34.51 ng/ml




#3 n-Heptane

R.T.: 7.997 min
Delta R.T.: -0.008 min
Response: 174920
Conc: 10.85 ng/ml




#4 Benzene

R.T.: 8.134 min
Delta R.T.: -0.007 min
Response: 194105
Conc: 10.29 ng/ml




#5 AAA-TFT

R.T.: 9.041 min
Delta R.T.: -0.008 min
Response: 199530
Conc: 17.07 ng/ml




#6 Toluene

R.T.: 10.888 min
Delta R.T.: -0.008 min
Response: 525635
Conc: 30.05 ng/ml




#7 Ethylbenzene

R.T.: 13.340 min
Delta R.T.: -0.009 min
Response: 208784
Conc: 9.51 ng/ml



#8 m-Xylene

R.T.: 13.473 min
Delta R.T.: -0.009 min
Response: 433135
Conc: 18.92 ng/ml




#9 O-Xylene

R.T.: 14.205 min

Delta R.T.: -0.009 min

Response: 433488

Conc: 19.41 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.491 min

Delta R.T.: -0.009 min

Response: 432946

Conc: 18.69 ng/ml