

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB022219\
 Data File : FB019878.D
 Signal(s) : FID2B.CH
 Acq On : 22 Feb 2019 15:30
 Operator : MJ/SJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Feb 22 22:46:20 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB012919.M
 Quant Title :
 QLast Update : Tue Jan 29 12:07:12 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	9.037	215750	18.457 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.893	587607	33.940 ng/ml
2) t 2,2,4-Trimethylpentane	7.652	576657	34.773 ng/ml
3) t n-Heptane	7.993	181350	11.254 ng/ml
4) t Benzene	8.130	203169	10.772 ng/ml
6) t Toluene	10.884	536620	30.677 ng/ml
7) t Ethylbenzene	13.337	215254	9.809 ng/ml
8) t m-Xylene	13.470	446180	19.494 ng/ml
9) t O-Xylene	14.201	432302	19.356 ng/ml
10) t 1,2,4-Trimethylbenzene	16.488	436930	18.857 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB022219\
Data File : FB019878.D
Signal(s) : FID2B.CH
Acq On : 22 Feb 2019 15:30
Operator : MJ/SJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 9 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Feb 22 22:46:20 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB012919.M
Quant Title :
QLast Update : Tue Jan 29 12:07:12 2019
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.893 min
Delta R.T.: -0.009 min
Response: 587607
Conc: 33.94 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.652 min
Delta R.T.: -0.012 min
Response: 576657
Conc: 34.77 ng/ml




#3 n-Heptane

R.T.: 7.993 min
Delta R.T.: -0.011 min
Response: 181350
Conc: 11.25 ng/ml




#4 Benzene

R.T.: 8.130 min
Delta R.T.: -0.011 min
Response: 203169
Conc: 10.77 ng/ml




#5 AAA-TFT

R.T.: 9.037 min
Delta R.T.: -0.012 min
Response: 215750
Conc: 18.46 ng/ml




#6 Toluene

R.T.: 10.884 min
Delta R.T.: -0.012 min
Response: 536620
Conc: 30.68 ng/ml




#7 Ethylbenzene

R.T.: 13.337 min
Delta R.T.: -0.012 min
Response: 215254
Conc: 9.81 ng/ml



#8 m-Xylene

R.T.: 13.470 min
Delta R.T.: -0.012 min
Response: 446180
Conc: 19.49 ng/ml




#9 O-Xylene

R.T.: 14.201 min

Delta R.T.: -0.013 min

Response: 432302

Conc: 19.36 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.488 min

Delta R.T.: -0.012 min

Response: 436930

Conc: 18.86 ng/ml