

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB022519\
 Data File : FB019886.D
 Signal(s) : FID2B.CH
 Acq On : 25 Feb 2019 12:35
 Operator : MJ/SJ
 Sample : K1586-01MS
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Feb 26 01:59:15 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB012919.M
 Quant Title :
 QLast Update : Tue Jan 29 12:07:12 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	9.038	236828	20.260 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.885	587109	33.911 ng/ml
2) t 2,2,4-Trimethylpentane	7.651	590242	35.592 ng/ml
3) t n-Heptane	7.991	150376	9.332 ng/ml
4) t Benzene	8.128	203482	10.788 ng/ml
6) t Toluene	10.886	524190	29.967 ng/ml
7) t Ethylbenzene	13.339	210594	9.596 ng/ml
8) t m-Xylene	13.473	440486	19.246 ng/ml
9) t O-Xylene	14.205	436865	19.560 ng/ml
10) t 1,2,4-Trimethylbenzene	16.492	432485	18.666 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB022519\
Data File : FB019886.D
Signal(s) : FID2B.CH
Acq On : 25 Feb 2019 12:35
Operator : MJ/SJ
Sample : K1586-01MS
Misc :
ALS Vial : 8 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Feb 26 01:59:15 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB012919.M
Quant Title :
QLast Update : Tue Jan 29 12:07:12 2019
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.885 min
Delta R.T.: -0.017 min
Response: 587109
Conc: 33.91 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.651 min
Delta R.T.: -0.014 min
Response: 590242
Conc: 35.59 ng/ml




#3 n-Heptane

R.T.: 7.991 min
Delta R.T.: -0.014 min
Response: 150376
Conc: 9.33 ng/ml




#4 Benzene

R.T.: 8.128 min
Delta R.T.: -0.013 min
Response: 203482
Conc: 10.79 ng/ml




#5 AAA-TFT

R.T.: 9.038 min
 Delta R.T.: -0.011 min
 Response: 236828
 Conc: 20.26 ng/ml




#6 Toluene

R.T.: 10.886 min
 Delta R.T.: -0.011 min
 Response: 524190
 Conc: 29.97 ng/ml




#7 Ethylbenzene

R.T.: 13.339 min
 Delta R.T.: -0.009 min
 Response: 210594
 Conc: 9.60 ng/ml




#8 m-Xylene

R.T.: 13.473 min
 Delta R.T.: -0.009 min
 Response: 440486
 Conc: 19.25 ng/ml



#9 O-Xylene

R.T.: 14.205 min
Delta R.T.: -0.009 min
Response: 436865
Conc: 19.56 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.492 min
Delta R.T.: -0.008 min
Response: 432485
Conc: 18.67 ng/ml