

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB022519\
 Data File : FB019886.D
 Signal(s) : FID2B.CH
 Acq On : 25 Feb 2019 12:35
 Operator : MJ/SJ
 Sample : K1586-01MS
 Misc : 5.04g/5.00ml DI WATER
 ALS Vial : 8 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Feb 25 12:54:38 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB012919.M
 Quant Title :
 QLast Update : Tue Jan 29 12:07:12 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	9.038	236828	21.175 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.885	587109	32.042 ng/ml
2) t 2,2,4-Trimethylpentane	7.651	590242	33.790 ng/ml
3) t n-Heptane	7.991	150376	9.033 ng/ml
4) t Benzene	8.128	203482	10.568 ng/ml
6) t Toluene	10.886	524190	30.118 ng/ml
7) t Ethylbenzene	13.339	210594	9.726 ng/ml
8) t m-Xylene	13.473	440486	19.503 ng/ml
9) t O-Xylene	14.205	436865	19.756 ng/ml
10) t 1,2,4-Trimethylbenzene	16.492	432485	19.001 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB022519\
Data File : FB019886.D
Signal(s) : FID2B.CH
Acq On : 25 Feb 2019 12:35
Operator : MJ/SJ
Sample : K1586-01MS
Misc : 5.04g/5.00ml DI WATER
ALS Vial : 8 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Feb 25 12:54:38 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB012919.M
Quant Title :
QLast Update : Tue Jan 29 12:07:12 2019
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.885 min
Delta R.T.: -0.017 min
Response: 587109
Conc: 32.04 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.651 min
Delta R.T.: -0.014 min
Response: 590242
Conc: 33.79 ng/ml




#3 n-Heptane

R.T.: 7.991 min
Delta R.T.: -0.014 min
Response: 150376
Conc: 9.03 ng/ml




#4 Benzene

R.T.: 8.128 min
Delta R.T.: -0.013 min
Response: 203482
Conc: 10.57 ng/ml




#5 AAA-TFT

R.T.: 9.038 min
Delta R.T.: -0.011 min
Response: 236828
Conc: 21.18 ng/ml




#6 Toluene

R.T.: 10.886 min
Delta R.T.: -0.011 min
Response: 524190
Conc: 30.12 ng/ml




#7 Ethylbenzene

R.T.: 13.339 min
Delta R.T.: -0.009 min
Response: 210594
Conc: 9.73 ng/ml




#8 m-Xylene

R.T.: 13.473 min
Delta R.T.: -0.009 min
Response: 440486
Conc: 19.50 ng/ml



#9 O-Xylene

R.T.: 14.205 min
Delta R.T.: -0.009 min
Response: 436865
Conc: 19.76 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.492 min
Delta R.T.: -0.008 min
Response: 432485
Conc: 19.00 ng/ml