

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB022619\
 Data File : FB019898.D
 Signal(s) : FID2B.CH
 Acq On : 26 Feb 2019 10:12
 Operator : MJ/SJ
 Sample : BSF0226S1
 Misc : 5.01g/5.00ml DI WATER
 ALS Vial : 3 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Feb 27 01:08:20 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB012919.M
 Quant Title :
 QLast Update : Tue Jan 29 12:07:12 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound		R.T.	Response	Conc Units

System Monitoring Compounds				
5) s	AAA-TFT	9.039	213322	18.249 ng/ml
Target Compounds				
1) t	2-Methylpentane	4.895	575747	33.255 ng/ml
2) t	2,2,4-Trimethylpentane	7.655	564989	34.069 ng/ml
3) t	n-Heptane	7.995	176436	10.949 ng/ml
4) t	Benzene	8.132	196460	10.416 ng/ml
6) t	Toluene	10.887	506456	28.953 ng/ml
7) t	Ethylbenzene	13.338	206859	9.426 ng/ml
8) t	m-Xylene	13.471	431969	18.873 ng/ml
9) t	O-Xylene	14.203	420829	18.842 ng/ml
10) t	1,2,4-Trimethylbenzene	16.489	433864	18.725 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB022619\
Data File : FB019898.D
Signal(s) : FID2B.CH
Acq On : 26 Feb 2019 10:12
Operator : MJ/SJ
Sample : BSF0226S1
Misc : 5.01g/5.00ml DI WATER
ALS Vial : 3 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Feb 27 01:08:20 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB012919.M
Quant Title :
QLast Update : Tue Jan 29 12:07:12 2019
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.895 min
Delta R.T.: -0.007 min
Response: 575747
Conc: 33.26 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.655 min
Delta R.T.: -0.009 min
Response: 564989
Conc: 34.07 ng/ml




#3 n-Heptane

R.T.: 7.995 min
Delta R.T.: -0.009 min
Response: 176436
Conc: 10.95 ng/ml




#4 Benzene

R.T.: 8.132 min
Delta R.T.: -0.009 min
Response: 196460
Conc: 10.42 ng/ml




#5 AAA-TFT

R.T.: 9.039 min
Delta R.T.: -0.010 min
Response: 213322
Conc: 18.25 ng/ml




#6 Toluene

R.T.: 10.887 min
Delta R.T.: -0.010 min
Response: 506456
Conc: 28.95 ng/ml




#7 Ethylbenzene

R.T.: 13.338 min
Delta R.T.: -0.010 min
Response: 206859
Conc: 9.43 ng/ml



#8 m-Xylene

R.T.: 13.471 min
Delta R.T.: -0.011 min
Response: 431969
Conc: 18.87 ng/ml




#9 O-Xylene

R.T.: 14.203 min

Delta R.T.: -0.010 min

Response: 420829

Conc: 18.84 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.489 min

Delta R.T.: -0.011 min

Response: 433864

Conc: 18.73 ng/ml