

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB022820\
 Data File : FB023750.D
 Signal(s) : FID2B.CH
 Acq On : 28 Feb 2020 18:10
 Operator : DD/AJ
 Sample : L1743-08
 Misc : 5.04g/5.00ml DI WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :

Integration File: Calibration.e
 Quant Time: Feb 28 22:54:15 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB020620.M
 Quant Title :
 QLast Update : Thu Feb 06 11:13:47 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.798	387376	19.027 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.684	1876	0.062 ng/ml
10) t 1,2,4-Trimethylbenzene	16.199	1193	0.041 ng/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB022820\
Data File : FB023750.D
Signal(s) : FID2B.CH
Acq On : 28 Feb 2020 18:10
Operator : DD/AJ
Sample : L1743-08
Misc : 5.04g/5.00ml DI WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :

Integration File: Calibration.e
Quant Time: Feb 28 22:54:15 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB020620.M
Quant Title :
QLast Update : Thu Feb 06 11:13:47 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane


R.T.: 4.684 min
Delta R.T.: -0.005 min
Response: 1876
Conc: 0.06 ng/ml

Instrument :
FID_B
ClientSampleId :



#5 AAA-TFT

R.T.: 8.798 min
Delta R.T.: 0.000 min
Response: 387376
Conc: 19.03 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.199 min
Delta R.T.: -0.010 min
Response: 1193
Conc: 0.04 ng/ml