

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB030220\
 Data File : FB023762.D
 Signal(s) : FID2B.CH
 Acq On : 2 Mar 2020 11:14
 Operator : DD/AJ
 Sample : L1743-04
 Misc : 5.02g/5.00ml DI WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :

Integration File: Calibration.e
 Quant Time: Mar 03 00:41:59 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB020620.M
 Quant Title :
 QLast Update : Thu Feb 06 11:13:47 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.798	433908	21.312 ng/ml
Target Compounds			
2) t 2,2,4-Trimethylpentane	7.382	165	0.005 ng/ml
8) t m-Xylene	13.210	1333	0.037 ng/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB030220\
Data File : FB023762.D
Signal(s) : FID2B.CH
Acq On : 2 Mar 2020 11:14
Operator : DD/AJ
Sample : L1743-04
Misc : 5.02g/5.00ml DI WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :

Integration File: Calibration.e
Quant Time: Mar 03 00:41:59 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB020620.M
Quant Title :
QLast Update : Thu Feb 06 11:13:47 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#2 2,2,4-Trimethylpentane


R.T.: 7.382 min
Delta R.T.: -0.036 min
Response: 165
Conc: 0.01 ng/ml

Instrument :
FID_B
ClientSampleId :



#5 AAA-TFT

R.T.: 8.798 min
Delta R.T.: 0.000 min
Response: 433908
Conc: 21.31 ng/ml



#8 m-Xylene

R.T.: 13.210 min
Delta R.T.: 0.007 min
Response: 1333
Conc: 0.04 ng/ml