

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB030221\
 Data File : FB025485.D
 Signal(s) : FID2B.CH
 Acq On : 2 Mar 2021 15:10
 Operator : DD/AJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :
 20 PPB GRO STD

Integration File: Calibration.e
 Quant Time: Mar 03 00:30:45 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB021921.M
 Quant Title :
 QLast Update : Mon Feb 22 11:51:04 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.802	560294	22.089 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.690	1035993	29.394 ng/ml
2) t 2,2,4-Trimethylpentane	7.420	1159434	31.225 ng/ml
3) t n-Heptane	7.760	379381	10.480 ng/ml
4) t Benzene	7.895	464244	10.049 ng/ml
6) t Toluene	10.633	1332398	29.801 ng/ml
7) t Ethylbenzene	13.074	410627	10.174 ng/ml
8) t m-Xylene	13.207	823114	20.461 ng/ml
9) t O-Xylene	13.934	785641	20.297 ng/ml
10) t 1,2,4-Trimethylbenzene	16.212	689434	20.538 ng/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB030221\
Data File : FB025485.D
Signal(s) : FID2B.CH
Acq On : 2 Mar 2021 15:10
Operator : DD/AJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :
20 PPB GRO STD

Integration File: Calibration.e
Quant Time: Mar 03 00:30:45 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB021921.M
Quant Title :
QLast Update : Mon Feb 22 11:51:04 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane


R.T.: 4.690 min
Delta R.T.: -0.002 min
Response: 1035993
Conc: 29.39 ng/ml

Instrument :
FID_B
ClientSampleId :
20 PPB GRO STD




#2 2,2,4-Trimethylpentane

R.T.: 7.420 min
Delta R.T.: -0.002 min
Response: 1159434
Conc: 31.22 ng/ml




#3 n-Heptane

R.T.: 7.760 min
Delta R.T.: -0.001 min
Response: 379381
Conc: 10.48 ng/ml



#4 Benzene


R.T.: 7.895 min
Delta R.T.: 0.000 min
Response: 464244
Conc: 10.05 ng/ml



#5 AAA-TFT


R.T.: 8.802 min
Delta R.T.: 0.000 min
Response: 560294
Conc: 22.09 ng/ml

Instrument :
FID_B
ClientSampleId :
20 PPB GRO STD




#6 Toluene

R.T.: 10.633 min
Delta R.T.: 0.000 min
Response: 1332398
Conc: 29.80 ng/ml




#7 Ethylbenzene

R.T.: 13.074 min
Delta R.T.: 0.000 min
Response: 410627
Conc: 10.17 ng/ml



#8 m-Xylene


R.T.: 13.207 min
Delta R.T.: 0.000 min
Response: 823114
Conc: 20.46 ng/ml



#9 O-Xylene

R.T.: 13.934 min
Delta R.T.: 0.000 min
Response: 785641
Conc: 20.30 ng/ml

Instrument :
FID_B
ClientSampleId :
20 PPB GRO STD



#10 1,2,4-Trimethylbenzene

R.T.: 16.212 min
Delta R.T.: 0.000 min
Response: 689434
Conc: 20.54 ng/ml