

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB030320\
 Data File : FB023768.D
 Signal(s) : FID2B.CH
 Acq On : 3 Mar 2020 9:21
 Operator : DD/AJ
 Sample : VBF0303S1
 Misc : 5.01g/5.00ml DI WATER
 ALS Vial : 2 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Mar 04 00:18:06 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB020620.M
 Quant Title :
 QLast Update : Thu Feb 06 11:13:47 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound		R.T.	Response	Conc Units

System Monitoring Compounds				
5) s	AAA-TFT	8.799	425537	20.901 ng/ml
Target Compounds				
1) t	2-Methylpentane	4.687	9365	0.311 ng/ml
2) t	2,2,4-Trimethylpentane	7.390	211	0.006 ng/ml
6) t	Toluene	10.632	4501	0.113 ng/ml
8) t	m-Xylene	13.202	2273	0.064 ng/ml
9) t	O-Xylene	13.938	1073	0.031 ng/ml
10) t	1,2,4-Trimethylbenzene	16.210	7913	0.270 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB030320\
Data File : FB023768.D
Signal(s) : FID2B.CH
Acq On : 3 Mar 2020 9:21
Operator : DD/AJ
Sample : VBF0303S1
Misc : 5.01g/5.00ml DI WATER
ALS Vial : 2 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Mar 04 00:18:06 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB020620.M
Quant Title :
QLast Update : Thu Feb 06 11:13:47 2020
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.687 min
Delta R.T.: -0.002 min
Response: 9365
Conc: 0.31 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.390 min
Delta R.T.: -0.028 min
Response: 211
Conc: 0.01 ng/ml




#5 AAA-TFT

R.T.: 8.799 min
Delta R.T.: 0.000 min
Response: 425537
Conc: 20.90 ng/ml




#6 Toluene

R.T.: 10.632 min
Delta R.T.: 0.002 min
Response: 4501
Conc: 0.11 ng/ml




#8 m-Xylene

R.T.: 13.202 min
Delta R.T.: 0.000 min
Response: 2273
Conc: 0.06 ng/ml



#9 O-Xylene

R.T.: 13.938 min
Delta R.T.: 0.007 min
Response: 1073
Conc: 0.03 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.210 min
Delta R.T.: 0.000 min
Response: 7913
Conc: 0.27 ng/ml