

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB030325\
 Data File : FB031525.D
 Signal(s) : FID2B.CH
 Acq On : 3 Mar 2025 11:06
 Operator : YP/AJ
 Sample : VBF0303S1
 Misc : 5.00G/5.00 ML DI WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :

Integration File: Calibration.e
 Quant Time: Mar 04 00:18:29 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB021125.M
 Quant Title :
 QLast Update : Tue Feb 11 12:33:57 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.787	522386	22.220 ng/ml
Target Compounds			
6) t Toluene	10.611	6521	0.159 ng/ml
8) t m-Xylene	13.187	5242	0.133 ng/ml
9) t O-Xylene	13.911	6755	0.181 ng/ml
10) t 1,2,4-Trimethylbenzene	16.190	13523	0.456 ng/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB030325\
Data File : FB031525.D
Signal(s) : FID2B.CH
Acq On : 3 Mar 2025 11:06
Operator : YP/AJ
Sample : VBF0303S1
Misc : 5.00G/5.00 ML DI WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :

Integration File: Calibration.e
Quant Time: Mar 04 00:18:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB021125.M
Quant Title :
QLast Update : Tue Feb 11 12:33:57 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#5 AAA-TFT


R.T.: 8.787 min
Delta R.T.: 0.001 min
Response: 522386
Conc: 22.22 ng/ml

Instrument :
FID_B
ClientSampleId :




#6 Toluene

R.T.: 10.611 min
Delta R.T.: -0.004 min
Response: 6521
Conc: 0.16 ng/ml




#8 m-Xylene

R.T.: 13.187 min
Delta R.T.: 0.001 min
Response: 5242
Conc: 0.13 ng/ml



#9 O-Xylene

R.T.: 13.911 min
Delta R.T.: -0.003 min
Response: 6755
Conc: 0.18 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.190 min
Delta R.T.: 0.000 min
Response: 13523
Conc: 0.46 ng/ml

Instrument :
FID_B
ClientSampleId :