

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB030325\
 Data File : FB031533.D
 Signal(s) : FID2B.CH
 Acq On : 3 Mar 2025 16:43
 Operator : YP/AJ
 Sample : BSF0303S2
 Misc : 5.00G/5.00 ML DI WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :
 BSF0303S2

Integration File: SAMPLE.e
 Quant Time: Mar 04 00:20:16 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB021125.M
 Quant Title :
 QLast Update : Tue Feb 11 12:33:57 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.790	493564	20.994 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.717	603153	18.882 ng/ml
2) t 2,2,4-Trimethylpentane	7.421	834820	21.455 ng/ml
3) t n-Heptane	7.752	257844	7.486 ng/ml
4) t Benzene	7.890	345829	7.789 ng/ml
6) t Toluene	10.619	993704	24.191 ng/ml
7) t Ethylbenzene	13.056	295228	8.132 ng/ml
8) t m-Xylene	13.190	647546	16.480 ng/ml
9) t O-Xylene	13.918	637911	17.055 ng/ml
10) t 1,2,4-Trimethylbenzene	16.194	486780	16.426 ng/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB030325\
Data File : FB031533.D
Signal(s) : FID2B.CH
Acq On : 3 Mar 2025 16:43
Operator : YP/AJ
Sample : BSF0303S2
Misc : 5.00G/5.00 ML DI WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :
BSF0303S2

Integration File: SAMPLE.e
Quant Time: Mar 04 00:20:16 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB021125.M
Quant Title :
QLast Update : Tue Feb 11 12:33:57 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um

