

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB030419\
Data File : FB020028.D
Signal(s) : FID2B.CH
Acq On : 4 Mar 2019 10:20
Operator : MJ/SJ
Sample : VBF0304S1
Misc : 5.02g/5.00ml DI WATER
ALS Vial : 2 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Mar 05 02:14:43 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB030119.M
Quant Title :
QLast Update : Fri Mar 01 06:28:25 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um

Compound		R.T.	Response	Conc Units

System Monitoring Compounds				
5) s	AAA-TFT	9.041	239773	21.397 ng/ml
Target Compounds				
2) t	2,2,4-Trimethylpentane	7.647	1522	0.076 ng/ml
6) t	Toluene	10.897	2554	0.135 ng/ml
8) t	m-Xylene	13.470	5411	0.231 ng/ml
10) t	1,2,4-Trimethylbenzene	16.478	6560	0.285 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB030419\
Data File : FB020028.D
Signal(s) : FID2B.CH
Acq On : 4 Mar 2019 10:20
Operator : MJ/SJ
Sample : VBF0304S1
Misc : 5.02g/5.00ml DI WATER
ALS Vial : 2 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Mar 05 02:14:43 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB030119.M
Quant Title :
QLast Update : Fri Mar 01 06:28:25 2019
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#2 2,2,4-Trimethylpentane

R.T.: 7.647 min
Delta R.T.: -0.004 min
Response: 1522
Conc: 0.08 ng/ml




#5 AAA-TFT

R.T.: 9.041 min
Delta R.T.: 0.005 min
Response: 239773
Conc: 21.40 ng/ml




#6 Toluene

R.T.: 10.897 min
Delta R.T.: 0.014 min
Response: 2554
Conc: 0.13 ng/ml



#8 m-Xylene

R.T.: 13.470 min
Delta R.T.: 0.000 min
Response: 5411
Conc: 0.23 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.478 min
Delta R.T.: -0.009 min
Response: 6560
Conc: 0.28 ng/ml