

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB030619\
 Data File : FB020079.D
 Signal(s) : FID2B.CH
 Acq On : 6 Mar 2019 18:20
 Operator : MJ/SJ
 Sample : K1753-02MS
 Misc : 5.04g/5.00ml DI WATER
 ALS Vial : 12 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Mar 07 08:06:33 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB030119.M
 Quant Title :
 QLast Update : Fri Mar 01 06:28:25 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	9.042	207951	18.557 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.906	441667	21.228 ng/ml
2) t 2,2,4-Trimethylpentane	7.673	411933	20.461 ng/ml
3) t n-Heptane	7.998	125016	6.703 ng/ml
4) t Benzene	8.135	176894	8.439 ng/ml
6) t Toluene	10.888	523952	27.646 ng/ml
7) t Ethylbenzene	13.340	196810	8.665 ng/ml
8) t m-Xylene	13.473	401301	17.115 ng/ml
9) t O-Xylene	14.205	390509	16.772 ng/ml
10) t 1,2,4-Trimethylbenzene	16.491	360525	15.659 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB030619\
Data File : FB020079.D
Signal(s) : FID2B.CH
Acq On : 6 Mar 2019 18:20
Operator : MJ/SJ
Sample : K1753-02MS
Misc : 5.04g/5.00ml DI WATER
ALS Vial : 12 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Mar 07 08:06:33 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB030119.M
Quant Title :
QLast Update : Fri Mar 01 06:28:25 2019
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.906 min
Delta R.T.: 0.015 min
Response: 441667
Conc: 21.23 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.673 min
Delta R.T.: 0.022 min
Response: 411933
Conc: 20.46 ng/ml




#3 n-Heptane

R.T.: 7.998 min
Delta R.T.: 0.007 min
Response: 125016
Conc: 6.70 ng/ml




#4 Benzene

R.T.: 8.135 min
Delta R.T.: 0.006 min
Response: 176894
Conc: 8.44 ng/ml




#5 AAA-TFT

R.T.: 9.042 min
Delta R.T.: 0.006 min
Response: 207951
Conc: 18.56 ng/ml




#6 Toluene

R.T.: 10.888 min
Delta R.T.: 0.005 min
Response: 523952
Conc: 27.65 ng/ml




#7 Ethylbenzene

R.T.: 13.340 min
Delta R.T.: 0.004 min
Response: 196810
Conc: 8.67 ng/ml




#8 m-Xylene

R.T.: 13.473 min
Delta R.T.: 0.004 min
Response: 401301
Conc: 17.11 ng/ml



#9 O-Xylene

R.T.: 14.205 min
Delta R.T.: 0.004 min
Response: 390509
Conc: 16.77 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.491 min
Delta R.T.: 0.004 min
Response: 360525
Conc: 15.66 ng/ml