

Method Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\
 Method File : FB030625.M
 Title :
 Last Update : Thu Mar 06 13:17:04 2025
 Response Via : Initial Calibration

Calibration Files

= S1 =FB031561.D S2 =FB031557.D
 S3 =FB031558.D S4 =FB031559.D S5 =FB031560.D

	Compound	S1	S2	S3	S4	S5	Avg	%RSD
1) t	2-Methylpentane	2.712	2.417	2.573	2.666	2.349	2.543 E4	6.15
2) t	2,2,4-Trimeth...	3.999	3.146	3.510	3.531	3.149	3.467 E4	10.13
3) t	n-Heptane	3.380	2.741	3.072	3.220	3.065	3.096 E4	7.64
4) t	Benzene	4.526	3.846	4.312	4.313	4.135	4.226 E4	6.00
5) s	AAA-TFT	1.762	2.310	2.411	2.405	2.435	2.264 E4	12.57
6) t	Toluene	4.502	3.502	3.951	4.009	3.815	3.956 E4	9.17
7) t	Ethylbenzene	4.211	3.131	3.492	3.497	3.304	3.527 E4	11.66
8) t	m-Xylene	4.612	3.408	3.820	3.829	3.605	3.855 E4	11.87
9) t	O-Xylene	4.353	3.440	3.673	3.618	3.369	3.690 E4	10.59
10) t	1,2,4-Trimeth...	3.734	2.597	2.810	2.641	2.345	2.825 E4	18.92

(#) = Out of Range