

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB030920\
Data File : FB023816.D
Signal(s) : FID2B.CH
Acq On : 9 Mar 2020 19:56
Operator : DD/AJ
Sample : L1852-06
Misc : 5.01g/5.00ml DI WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :

Integration File: Calibration.e
Quant Time: Mar 10 13:47:48 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB030520.M
Quant Title :
QLast Update : Thu Mar 05 11:03:12 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.800	347462	16.506 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.688	1284	0.042 ng/ml
3) t n-Heptane	7.768	1033	0.031 ng/ml
6) t Toluene	10.695	1790	0.044 ng/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB030920\
Data File : FB023816.D
Signal(s) : FID2B.CH
Acq On : 9 Mar 2020 19:56
Operator : DD/AJ
Sample : L1852-06
Misc : 5.01g/5.00ml DI WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :

Integration File: Calibration.e
Quant Time: Mar 10 13:47:48 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB030520.M
Quant Title :
QLast Update : Thu Mar 05 11:03:12 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane


R.T.: 4.688 min
Delta R.T.: 0.002 min
Response: 1284
Conc: 0.04 ng/ml

Instrument :
FID_B
ClientSampleId :




#3 n-Heptane

R.T.: 7.768 min
Delta R.T.: 0.015 min
Response: 1033
Conc: 0.03 ng/ml



#5 AAA-TFT

R.T.: 8.800 min
Delta R.T.: 0.006 min
Response: 347462
Conc: 16.51 ng/ml



#6 Toluene

R.T.: 10.695 min
Delta R.T.: 0.069 min
Response: 1790
Conc: 0.04 ng/ml