

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB031020\
 Data File : FB023858.D
 Signal(s) : FID2B.CH
 Acq On : 11 Mar 2020 6:45
 Operator : DD/AJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Mar 11 07:11:59 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB030520.M
 Quant Title :
 QLast Update : Thu Mar 05 11:03:12 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.798	455440	21.635 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.689	970068	31.389 ng/ml
2) t 2,2,4-Trimethylpentane	7.416	1093550	32.430 ng/ml
3) t n-Heptane	7.757	352747	10.648 ng/ml
4) t Benzene	7.891	431752	10.311 ng/ml
6) t Toluene	10.630	1266213	31.055 ng/ml
7) t Ethylbenzene	13.070	380535	9.960 ng/ml
8) t m-Xylene	13.203	731927	19.880 ng/ml
9) t O-Xylene	13.930	707864	19.903 ng/ml
10) t 1,2,4-Trimethylbenzene	16.208	603826	19.931 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB031020\
Data File : FB023858.D
Signal(s) : FID2B.CH
Acq On : 11 Mar 2020 6:45
Operator : DD/AJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 32 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Mar 11 07:11:59 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB030520.M
Quant Title :
QLast Update : Thu Mar 05 11:03:12 2020
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.689 min
Delta R.T.: 0.003 min
Response: 970068
Conc: 31.39 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.416 min
Delta R.T.: 0.003 min
Response: 1093550
Conc: 32.43 ng/ml




#3 n-Heptane

R.T.: 7.757 min
Delta R.T.: 0.003 min
Response: 352747
Conc: 10.65 ng/ml




#4 Benzene

R.T.: 7.891 min
Delta R.T.: 0.003 min
Response: 431752
Conc: 10.31 ng/ml




#5 AAA-TFT

R.T.: 8.798 min
Delta R.T.: 0.004 min
Response: 455440
Conc: 21.64 ng/ml




#6 Toluene

R.T.: 10.630 min
Delta R.T.: 0.003 min
Response: 1266213
Conc: 31.05 ng/ml




#7 Ethylbenzene

R.T.: 13.070 min
Delta R.T.: 0.003 min
Response: 380535
Conc: 9.96 ng/ml




#8 m-Xylene

R.T.: 13.203 min
Delta R.T.: 0.003 min
Response: 731927
Conc: 19.88 ng/ml



#9 O-Xylene

R.T.: 13.930 min
Delta R.T.: 0.003 min
Response: 707864
Conc: 19.90 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.208 min
Delta R.T.: 0.002 min
Response: 603826
Conc: 19.93 ng/ml