

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB031219\
 Data File : FB020155.D
 Signal(s) : FID2B.CH
 Acq On : 12 Mar 2019 12:35
 Operator : MJ/SJ
 Sample : K1846-06
 Misc : 5.01g/5.00ml DI WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :
 TP-1-S2

Integration File: Calibration.e
 Quant Time: Mar 13 02:28:06 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB031119.M
 Quant Title :
 QLast Update : Mon Mar 11 13:44:46 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	9.041	482175	18.367 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.896	4013	0.119 ng/ml
2) t 2,2,4-Trimethylpentane	7.678	17110	0.522 ng/ml
6) t Toluene	10.884	4457	0.121 ng/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB031219\
Data File : FB020155.D
Signal(s) : FID2B.CH
Acq On : 12 Mar 2019 12:35
Operator : MJ/SJ
Sample : K1846-06
Misc : 5.01g/5.00ml DI WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :
TP-1-S2

Integration File: Calibration.e
Quant Time: Mar 13 02:28:06 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB031119.M
Quant Title :
QLast Update : Mon Mar 11 13:44:46 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane


R.T.: 4.896 min
Delta R.T.: 0.002 min
Response: 4013
Conc: 0.12 ng/ml

Instrument :
FID_B
ClientSampleId :
TP-1-S2




#2 2,2,4-Trimethylpentane

R.T.: 7.678 min
Delta R.T.: 0.021 min
Response: 17110
Conc: 0.52 ng/ml



#5 AAA-TFT

R.T.: 9.041 min
Delta R.T.: 0.003 min
Response: 482175
Conc: 18.37 ng/ml



#6 Toluene

R.T.: 10.884 min
Delta R.T.: -0.002 min
Response: 4457
Conc: 0.12 ng/ml