

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB031720\
 Data File : FB023972.D
 Signal(s) : FID2B.CH
 Acq On : 17 Mar 2020 13:40
 Operator : DD/AJ
 Sample : VBF0317S2
 Misc : 5.01g/5.00ml DI WATER
 ALS Vial : 8 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Mar 18 00:49:08 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB030520.M
 Quant Title :
 QLast Update : Thu Mar 05 11:03:12 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.797	448838	21.322 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.692	1786	0.058 ng/ml
2) t 2,2,4-Trimethylpentane	7.398	1007	0.030 ng/ml
3) t n-Heptane	7.775	2268	0.068 ng/ml
6) t Toluene	10.618	1228	0.030 ng/ml
8) t m-Xylene	13.194	7352	0.200 ng/ml
9) t O-Xylene	13.924	5029	0.141 ng/ml
10) t 1,2,4-Trimethylbenzene	16.223	3164	0.104 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB031720\
Data File : FB023972.D
Signal(s) : FID2B.CH
Acq On : 17 Mar 2020 13:40
Operator : DD/AJ
Sample : VBF0317S2
Misc : 5.01g/5.00ml DI WATER
ALS Vial : 8 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Mar 18 00:49:08 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB030520.M
Quant Title :
QLast Update : Thu Mar 05 11:03:12 2020
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.692 min
Delta R.T.: 0.006 min
Response: 1786
Conc: 0.06 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.398 min
Delta R.T.: -0.015 min
Response: 1007
Conc: 0.03 ng/ml




#3 n-Heptane

R.T.: 7.775 min
Delta R.T.: 0.021 min
Response: 2268
Conc: 0.07 ng/ml




#5 AAA-TFT

R.T.: 8.797 min
Delta R.T.: 0.003 min
Response: 448838
Conc: 21.32 ng/ml




#6 Toluene

R.T.: 10.618 min
Delta R.T.: -0.009 min
Response: 1228
Conc: 0.03 ng/ml




#8 m-Xylene

R.T.: 13.194 min
Delta R.T.: -0.006 min
Response: 7352
Conc: 0.20 ng/ml



#9 o-Xylene

R.T.: 13.924 min
Delta R.T.: -0.004 min
Response: 5029
Conc: 0.14 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.223 min
Delta R.T.: 0.017 min
Response: 3164
Conc: 0.10 ng/ml