

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB031821\
 Data File : FB025544.D
 Signal(s) : FID2B.CH
 Acq On : 18 Mar 2021 13:34
 Operator : DD/AJ
 Sample : BSF0318S1
 Misc : 5.01g/5.00ml DI WATER
 ALS Vial : 3 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Mar 19 01:06:08 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB031621.M
 Quant Title :
 QLast Update : Wed Mar 17 08:44:23 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.803	467975	22.356 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.691	1029363	34.546 ng/ml
2) t 2,2,4-Trimethylpentane	7.420	1134950	35.781 ng/ml
3) t n-Heptane	7.761	369286	11.897 ng/ml
4) t Benzene	7.896	445745	11.471 ng/ml
6) t Toluene	10.634	1289539	34.151 ng/ml
7) t Ethylbenzene	13.074	398673	11.540 ng/ml
8) t m-Xylene	13.207	791144	23.064 ng/ml
9) t O-Xylene	13.934	757524	22.875 ng/ml
10) t 1,2,4-Trimethylbenzene	16.212	670498	23.234 ng/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB031821\
Data File : FB025544.D
Signal(s) : FID2B.CH
Acq On : 18 Mar 2021 13:34
Operator : DD/AJ
Sample : BSF0318S1
Misc : 5.01g/5.00ml DI WATER
ALS Vial : 3 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Mar 19 01:06:08 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB031621.M
Quant Title :
QLast Update : Wed Mar 17 08:44:23 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um

