

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB031920\
 Data File : FB024000.D
 Signal(s) : FID2B.CH
 Acq On : 19 Mar 2020 17:10
 Operator : DD/AJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Mar 20 00:58:01 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB030520.M
 Quant Title :
 QLast Update : Thu Mar 05 11:03:12 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.799	465167	22.097 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.689	920721	29.792 ng/ml
2) t 2,2,4-Trimethylpentane	7.417	1060977	31.464 ng/ml
3) t n-Heptane	7.757	337185	10.178 ng/ml
4) t Benzene	7.892	420145	10.034 ng/ml
6) t Toluene	10.630	1257249	30.835 ng/ml
7) t Ethylbenzene	13.070	391064	10.236 ng/ml
8) t m-Xylene	13.203	754809	20.502 ng/ml
9) t O-Xylene	13.930	723588	20.345 ng/ml
10) t 1,2,4-Trimethylbenzene	16.208	620687	20.488 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB031920\
Data File : FB024000.D
Signal(s) : FID2B.CH
Acq On : 19 Mar 2020 17:10
Operator : DD/AJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 13 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Mar 20 00:58:01 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB030520.M
Quant Title :
QLast Update : Thu Mar 05 11:03:12 2020
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.689 min
Delta R.T.: 0.003 min
Response: 920721
Conc: 29.79 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.417 min
Delta R.T.: 0.004 min
Response: 1060977
Conc: 31.46 ng/ml




#3 n-Heptane

R.T.: 7.757 min
Delta R.T.: 0.004 min
Response: 337185
Conc: 10.18 ng/ml




#4 Benzene

R.T.: 7.892 min
Delta R.T.: 0.004 min
Response: 420145
Conc: 10.03 ng/ml




#5 AAA-TFT

R.T.: 8.799 min
Delta R.T.: 0.004 min
Response: 465167
Conc: 22.10 ng/ml




#6 Toluene

R.T.: 10.630 min
Delta R.T.: 0.004 min
Response: 1257249
Conc: 30.83 ng/ml




#7 Ethylbenzene

R.T.: 13.070 min
Delta R.T.: 0.003 min
Response: 391064
Conc: 10.24 ng/ml




#8 m-Xylene

R.T.: 13.203 min
Delta R.T.: 0.003 min
Response: 754809
Conc: 20.50 ng/ml



#9 O-Xylene

R.T.: 13.930 min
Delta R.T.: 0.003 min
Response: 723588
Conc: 20.35 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.208 min
Delta R.T.: 0.002 min
Response: 620687
Conc: 20.49 ng/ml