

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB032020\
 Data File : FB024014.D
 Signal(s) : FID2B.CH
 Acq On : 20 Mar 2020 21:12
 Operator : DD/AJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Mar 20 23:03:43 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB030520.M
 Quant Title :
 QLast Update : Thu Mar 05 11:03:12 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound		R.T.	Response	Conc Units

System Monitoring Compounds				
5) s	AAA-TFT	8.804	431335	20.490 ng/ml
Target Compounds				
1) t	2-Methylpentane	4.694	911754	29.502 ng/ml
2) t	2,2,4-Trimethylpentane	7.423	1055956	31.315 ng/ml
3) t	n-Heptane	7.762	330719	9.983 ng/ml
4) t	Benzene	7.897	431035	10.294 ng/ml
6) t	Toluene	10.635	1244086	30.512 ng/ml
7) t	Ethylbenzene	13.076	377480	9.880 ng/ml
8) t	m-Xylene	13.209	727089	19.749 ng/ml
9) t	O-Xylene	13.936	710081	19.965 ng/ml
10) t	1,2,4-Trimethylbenzene	16.213	616484	20.349 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB032020\
Data File : FB024014.D
Signal(s) : FID2B.CH
Acq On : 20 Mar 2020 21:12
Operator : DD/AJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 15 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Mar 20 23:03:43 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB030520.M
Quant Title :
QLast Update : Thu Mar 05 11:03:12 2020
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.694 min
Delta R.T.: 0.008 min
Response: 911754
Conc: 29.50 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.423 min
Delta R.T.: 0.010 min
Response: 1055956
Conc: 31.31 ng/ml




#3 n-Heptane

R.T.: 7.762 min
Delta R.T.: 0.009 min
Response: 330719
Conc: 9.98 ng/ml




#4 Benzene

R.T.: 7.897 min
Delta R.T.: 0.009 min
Response: 431035
Conc: 10.29 ng/ml




#5 AAA-TFT

R.T.: 8.804 min
Delta R.T.: 0.010 min
Response: 431335
Conc: 20.49 ng/ml




#6 Toluene

R.T.: 10.635 min
Delta R.T.: 0.009 min
Response: 1244086
Conc: 30.51 ng/ml




#7 Ethylbenzene

R.T.: 13.076 min
Delta R.T.: 0.009 min
Response: 377480
Conc: 9.88 ng/ml




#8 m-Xylene

R.T.: 13.209 min
Delta R.T.: 0.009 min
Response: 727089
Conc: 19.75 ng/ml



#9 O-Xylene

R.T.: 13.936 min
Delta R.T.: 0.008 min
Response: 710081
Conc: 19.97 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.213 min
Delta R.T.: 0.008 min
Response: 616484
Conc: 20.35 ng/ml