

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB032023\
 Data File : FB029527.D
 Signal(s) : FID2B.CH
 Acq On : 20 Mar 2023 15:00
 Operator : YP/AJ
 Sample : 01950-18
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Integration File: SAMPLE.e
 Quant Time: Apr 05 07:50:08 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB030123.M
 Quant Title :
 QLast Update : Wed Mar 01 12:30:24 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.837	828912	68.417 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.701	186689	12.167 ng/ml
2) t 2,2,4-Trimethylpentane	7.391	620652	36.845 ng/ml
3) t n-Heptane	7.745	899041	51.980 ng/ml
4) t Benzene	7.874	202979	10.000 ng/ml
6) t Toluene	10.617	2697920	136.680 ng/ml
7) t Ethylbenzene	13.054	843805	45.324 ng/ml
8) t m-Xylene	13.186	2517279	135.312 ng/ml
9) t O-Xylene	13.915	987418	53.626 ng/ml
10) t 1,2,4-Trimethylbenzene	16.190	1010351	54.245 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB032023\
Data File : FB029527.D
Signal(s) : FID2B.CH
Acq On : 20 Mar 2023 15:00
Operator : YP/AJ
Sample : 01950-18
Misc :
ALS Vial : 11 Sample Multiplier: 1

Integration File: SAMPLE.e
Quant Time: Apr 05 07:50:08 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB030123.M
Quant Title :
QLast Update : Wed Mar 01 12:30:24 2023
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.701 min
Delta R.T.: -0.009 min
Response: 186689
Conc: 12.17 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.391 min
Delta R.T.: -0.025 min
Response: 620652
Conc: 36.85 ng/ml




#3 n-Heptane

R.T.: 7.745 min
Delta R.T.: -0.001 min
Response: 899041
Conc: 51.98 ng/ml




#4 Benzene

R.T.: 7.874 min
Delta R.T.: -0.009 min
Response: 202979
Conc: 10.00 ng/ml




#5 AAA-TFT

R.T.: 8.837 min
Delta R.T.: 0.050 min
Response: 828912
Conc: 68.42 ng/ml




#6 Toluene

R.T.: 10.617 min
Delta R.T.: 0.001 min
Response: 2697920
Conc: 136.68 ng/ml




#7 Ethylbenzene

R.T.: 13.054 min
Delta R.T.: 0.001 min
Response: 843805
Conc: 45.32 ng/ml




#8 m-Xylene

R.T.: 13.186 min
Delta R.T.: 0.000 min
Response: 2517279
Conc: 135.31 ng/ml



#9 O-Xylene

R.T.: 13.915 min
Delta R.T.: 0.003 min
Response: 987418
Conc: 53.63 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.190 min
Delta R.T.: 0.004 min
Response: 1010351
Conc: 54.24 ng/ml