

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB032219\
 Data File : FB020350.D
 Signal(s) : FID2B.CH
 Acq On : 22 Mar 2019 19:10
 Operator : MJ/SJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Mar 23 00:26:58 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB031119.M
 Quant Title :
 QLast Update : Mon Mar 11 13:44:46 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	9.045	551266	20.998 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.901	1035078	30.684 ng/ml
2) t 2,2,4-Trimethylpentane	7.668	914589	27.889 ng/ml
3) t n-Heptane	8.000	428384	10.293 ng/ml
4) t Benzene	8.137	480317	10.317 ng/ml
6) t Toluene	10.893	1087637	29.433 ng/ml
7) t Ethylbenzene	13.346	237897	10.111 ng/ml
8) t m-Xylene	13.479	444984	20.860 ng/ml
9) t O-Xylene	14.212	420237	22.184 ng/ml
10) t 1,2,4-Trimethylbenzene	16.498	443549	22.365 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB032219\
Data File : FB020350.D
Signal(s) : FID2B.CH
Acq On : 22 Mar 2019 19:10
Operator : MJ/SJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 13 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Mar 23 00:26:58 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB031119.M
Quant Title :
QLast Update : Mon Mar 11 13:44:46 2019
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.901 min
Delta R.T.: 0.007 min
Response: 1035078
Conc: 30.68 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.668 min
Delta R.T.: 0.011 min
Response: 914589
Conc: 27.89 ng/ml




#3 n-Heptane

R.T.: 8.000 min
Delta R.T.: 0.007 min
Response: 428384
Conc: 10.29 ng/ml




#4 Benzene

R.T.: 8.137 min
Delta R.T.: 0.007 min
Response: 480317
Conc: 10.32 ng/ml




#5 AAA-TFT

R.T.: 9.045 min
Delta R.T.: 0.007 min
Response: 551266
Conc: 21.00 ng/ml




#6 Toluene

R.T.: 10.893 min
Delta R.T.: 0.007 min
Response: 1087637
Conc: 29.43 ng/ml




#7 Ethylbenzene

R.T.: 13.346 min
Delta R.T.: 0.008 min
Response: 237897
Conc: 10.11 ng/ml




#8 m-Xylene

R.T.: 13.479 min
Delta R.T.: 0.008 min
Response: 444984
Conc: 20.86 ng/ml



#9 O-Xylene

R.T.: 14.212 min
Delta R.T.: 0.008 min
Response: 420237
Conc: 22.18 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.498 min
Delta R.T.: 0.009 min
Response: 443549
Conc: 22.37 ng/ml