

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB032222\
 Data File : FB026767.D
 Signal(s) : FID2B.CH
 Acq On : 22 Mar 2022 20:03
 Operator : AJ/MA
 Sample : N1974-06MS
 Misc : 5.01G/5.00 ML DI WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :
 G-6-4-6-7MS

Integration File: Calibration.e
 Quant Time: Mar 23 02:38:05 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB030822.M
 Quant Title :
 QLast Update : Fri Mar 11 13:08:49 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.789	239341	22.809 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.705	438493	23.160 ng/ml
2) t 2,2,4-Trimethylpentane	7.429	504444	24.337 ng/ml
3) t n-Heptane	7.746	153200	7.740 ng/ml
4) t Benzene	7.883	217109	9.409 ng/ml
6) t Toluene	10.616	616817	28.174 ng/ml
7) t Ethylbenzene	13.053	191616	9.534 ng/ml
8) t m-Xylene	13.186	381385	18.925 ng/ml
9) t O-Xylene	13.913	374642	18.779 ng/ml
10) t 1,2,4-Trimethylbenzene	16.189	379138	19.029 ng/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB032222\
Data File : FB026767.D
Signal(s) : FID2B.CH
Acq On : 22 Mar 2022 20:03
Operator : AJ/MA
Sample : N1974-06MS
Misc : 5.01G/5.00 ML DI WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :
G-6-4-6-7MS

Integration File: Calibration.e
Quant Time: Mar 23 02:38:05 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB030822.M
Quant Title :
QLast Update : Fri Mar 11 13:08:49 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um

