

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB032322\
 Data File : FB026788.D
 Signal(s) : FID2B.CH
 Acq On : 23 Mar 2022 11:47
 Operator : AJ/MA
 Sample : BSF0322S1
 Misc : 5.02G/5.00 ML DI WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :
 BSF0323S1

Integration File: Calibration.e
 Quant Time: Mar 23 16:08:16 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB030822.M
 Quant Title :
 QLast Update : Fri Mar 11 13:08:49 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.787	249275	23.755 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.704	439233	23.199 ng/ml
2) t 2,2,4-Trimethylpentane	7.427	514677	24.831 ng/ml
3) t n-Heptane	7.747	158724	8.019 ng/ml
4) t Benzene	7.882	220198	9.543 ng/ml
6) t Toluene	10.615	629676	28.762 ng/ml
7) t Ethylbenzene	13.052	196850	9.795 ng/ml
8) t m-Xylene	13.185	392563	19.479 ng/ml
9) t O-Xylene	13.912	382646	19.180 ng/ml
10) t 1,2,4-Trimethylbenzene	16.188	387949	19.471 ng/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB032322\
Data File : FB026788.D
Signal(s) : FID2B.CH
Acq On : 23 Mar 2022 11:47
Operator : AJ/MA
Sample : BSF0322S1
Misc : 5.02G/5.00 ML DI WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :
BSF0323S1

Integration File: Calibration.e
Quant Time: Mar 23 16:08:16 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB030822.M
Quant Title :
QLast Update : Fri Mar 11 13:08:49 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um

