

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE040523\
 Data File : FE041185.D
 Signal(s) : FID1B.ch
 Acq On : 05 Apr 2023 20:19
 Operator : YP\AJ
 Sample : FE040523ICV
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Apr 06 02:06:35 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE040523.M
 Quant Title :
 QLast Update : Thu Apr 06 02:04:41 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1	N-OCTANE	149.744	133.465 E3	10.9	92	0.00
2	N-DECANE	148.307	132.909 E3	10.4	91	0.00
3	N-DODECANE	151.729	135.492 E3	10.7	91	0.00
4	N-TETRADECANE	152.602	134.940 E3	11.6	91	0.00
5	N-HEXADECANE	154.023	135.565 E3	12.0	91	0.00
6	N-OCTADECANE	152.976	134.116 E3	12.3	91	0.00
7	N-EICOSANE	150.816	132.782 E3	12.0	91	0.00
8	N-DOCOSANE	144.598	126.585 E3	12.5	91	0.00
9 S	TETRACOSANE-d50 (SURROGATE)	127.279	110.967 E3	12.8	91	0.00
10	N-TETRACOSANE	142.816	124.585 E3	12.8	91	0.00
11	N-HEXACOSANE	139.110	121.155 E3	12.9	91	0.00
12	N-OCTACOSANE	137.377	119.126 E3	13.3	91	0.00
13	N-TRIACONTANE	136.607	116.762 E3	14.5	90	0.00
14	N-DOTRIACONTANE	133.279	112.317 E3	15.7	90	0.00
15	N-TETRATRIACONTANE	126.584	107.085 E3	15.4	90	0.00
16	N-HEXATRIACONTANE	120.734	104.359 E3	13.6	90	0.00
17	N-OCTATRIACONTANE	123.662	106.715 E3	13.7	91	0.00
18	N-TETRACONTANE	130.068	112.318 E3	13.6	90	0.00

Evaluate Continuing Calibration Report - Not Found

(#) = Out of Range

SPCC's out = 0 CCC's out = 0