

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB032619\
 Data File : FB020368.D
 Signal(s) : FID2B.CH
 Acq On : 26 Mar 2019 21:43
 Operator : MJ/SJ
 Sample : K2091-01MS
 Misc : 5.03gm/5.00ml DI WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :
 DRUM-COMPMS

Integration File: Calibration.e
 Quant Time: Mar 27 02:18:29 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB032619.M
 Quant Title :
 QLast Update : Tue Mar 26 12:48:09 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	9.040	280696	14.569 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.900	358153	12.921 ng/ml
2) t 2,2,4-Trimethylpentane	7.667	198290	8.269 ng/ml
3) t n-Heptane	7.996	90190	2.759 ng/ml
4) t Benzene	8.134	287557	8.303 ng/ml
6) t Toluene	10.887	561807	21.661 ng/ml
7) t Ethylbenzene	13.339	124515	6.325 ng/ml
8) t m-Xylene	13.472	258617	12.738 ng/ml
9) t O-Xylene	14.204	257858	12.732 ng/ml
10) t 1,2,4-Trimethylbenzene	16.490	212156	10.197 ng/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB032619\
Data File : FB020368.D
Signal(s) : FID2B.CH
Acq On : 26 Mar 2019 21:43
Operator : MJ/SJ
Sample : K2091-01MS
Misc : 5.03gm/5.00ml DI WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :
DRUM-COMPMS

Integration File: Calibration.e
Quant Time: Mar 27 02:18:29 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB032619.M
Quant Title :
QLast Update : Tue Mar 26 12:48:09 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um

