

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB040320\
Data File : FB024066.D
Signal(s) : FID2B.CH
Acq On : 3 Apr 2020 17:51
Operator : DD/AJ
Sample : L2122-02
Misc : 5.06g/5.00ml DI WATER
ALS Vial : 9 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Apr 04 00:49:57 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB032620.M
Quant Title :
QLast Update : Fri Mar 27 11:16:49 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.802	393663	17.235 ng/ml
Target Compounds			
3) t n-Heptane	7.728	595	0.016 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB040320\
Data File : FB024066.D
Signal(s) : FID2B.CH
Acq On : 3 Apr 2020 17:51
Operator : DD/AJ
Sample : L2122-02
Misc : 5.06g/5.00ml DI WATER
ALS Vial : 9 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Apr 04 00:49:57 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB032620.M
Quant Title :
QLast Update : Fri Mar 27 11:16:49 2020
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um





#3 n-Heptane

R.T.: 7.728 min
Delta R.T.: -0.032 min
Response: 595
Conc: 0.02 ng/ml



#5 AAA-TFT

R.T.: 8.802 min
Delta R.T.: 0.000 min
Response: 393663
Conc: 17.24 ng/ml