

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB040419\
 Data File : FB020532.D
 Signal(s) : FID2B.CH
 Acq On : 4 Apr 2019 17:27
 Operator : MJ/SJ
 Sample : FB040419GROICV
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :
 FB040419GROICV

Integration File: Calibration.e
 Quant Time: Apr 04 17:46:23 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB040419.M
 Quant Title :
 QLast Update : Thu Apr 04 17:27:15 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.999	360842	19.564 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.868	1094350	31.086 ng/ml
2) t 2,2,4-Trimethylpentane	7.628	928867	31.156 ng/ml
3) t n-Heptane	7.955	411334	10.478 ng/ml
4) t Benzene	8.091	424190	10.327 ng/ml
6) t Toluene	10.843	929330	30.756 ng/ml
7) t Ethylbenzene	13.292	259000	10.124 ng/ml
8) t m-Xylene	13.425	534905	20.239 ng/ml
9) t O-Xylene	14.157	549676	20.304 ng/ml
10) t 1,2,4-Trimethylbenzene	16.441	542095	20.748 ng/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB040419\
Data File : FB020532.D
Signal(s) : FID2B.CH
Acq On : 4 Apr 2019 17:27
Operator : MJ/SJ
Sample : FB040419GROICV
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :
FB040419GROICV

Integration File: Calibration.e
Quant Time: Apr 04 17:46:23 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB040419.M
Quant Title :
QLast Update : Thu Apr 04 17:27:15 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane


R.T.: 4.868 min
Delta R.T.: -0.001 min
Response: 1094350
Conc: 31.09 ng/ml

Instrument :
FID_B
ClientSampleId :
FB040419GROICV




#2 2,2,4-Trimethylpentane

R.T.: 7.628 min
Delta R.T.: 0.002 min
Response: 928867
Conc: 31.16 ng/ml




#3 n-Heptane

R.T.: 7.955 min
Delta R.T.: 0.000 min
Response: 411334
Conc: 10.48 ng/ml



#4 Benzene


R.T.: 8.091 min
Delta R.T.: 0.000 min
Response: 424190
Conc: 10.33 ng/ml



#5 AAA-TFT


R.T.: 8.999 min
Delta R.T.: 0.000 min
Response: 360842
Conc: 19.56 ng/ml

Instrument :
FID_B
ClientSampleId :
FB040419GROICV




#6 Toluene

R.T.: 10.843 min
Delta R.T.: 0.000 min
Response: 929330
Conc: 30.76 ng/ml




#7 Ethylbenzene

R.T.: 13.292 min
Delta R.T.: 0.000 min
Response: 259000
Conc: 10.12 ng/ml



#8 m-Xylene


R.T.: 13.425 min
Delta R.T.: 0.000 min
Response: 534905
Conc: 20.24 ng/ml



#9 O-Xylene

R.T.: 14.157 min
Delta R.T.: 0.000 min
Response: 549676
Conc: 20.30 ng/ml

Instrument :
FID_B
ClientSampleId :
FB040419GROICV



#10 1,2,4-Trimethylbenzene

R.T.: 16.441 min
Delta R.T.: 0.000 min
Response: 542095
Conc: 20.75 ng/ml