

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB040519\
 Data File : FB020539.D
 Signal(s) : FID2B.CH
 Acq On : 5 Apr 2019 15:00
 Operator : MJ/SJ
 Sample : K1825-18 5X
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :
 RR-GAS-WP

Integration File: Calibration.e
 Quant Time: Apr 06 01:17:10 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB040419.M
 Quant Title :
 QLast Update : Thu Apr 04 17:27:15 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	9.000	443072	24.022 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.865	120234	3.415 ng/ml
2) t 2,2,4-Trimethylpentane	7.581	150385	5.044 ng/ml
3) t n-Heptane	7.952	300929	7.666 ng/ml
4) t Benzene	8.088	117378	2.858 ng/ml
6) t Toluene	10.840	865982	28.660 ng/ml
7) t Ethylbenzene	13.290	195439	7.640 ng/ml
8) t m-Xylene	13.421	816548	30.896 ng/ml
9) t O-Xylene	14.155	326325	12.054 ng/ml
10) t 1,2,4-Trimethylbenzene	16.438	364558	13.953 ng/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB040519\
Data File : FB020539.D
Signal(s) : FID2B.CH
Acq On : 5 Apr 2019 15:00
Operator : MJ/SJ
Sample : K1825-18 5X
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :
RR-GAS-WP

Integration File: Calibration.e
Quant Time: Apr 06 01:17:10 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB040419.M
Quant Title :
QLast Update : Thu Apr 04 17:27:15 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane


R.T.: 4.865 min
Delta R.T.: -0.004 min
Response: 120234
Conc: 3.42 ng/ml

Instrument :
FID_B
ClientSampleId :
RR-GAS-WP




#2 2,2,4-Trimethylpentane

R.T.: 7.581 min
Delta R.T.: -0.045 min
Response: 150385
Conc: 5.04 ng/ml




#3 n-Heptane

R.T.: 7.952 min
Delta R.T.: -0.003 min
Response: 300929
Conc: 7.67 ng/ml



#4 Benzene


R.T.: 8.088 min
Delta R.T.: -0.003 min
Response: 117378
Conc: 2.86 ng/ml



#5 AAA-TFT


R.T.: 9.000 min
Delta R.T.: 0.001 min
Response: 443072
Conc: 24.02 ng/ml

Instrument :
FID_B
ClientSampleId :
RR-GAS-WP




#6 Toluene

R.T.: 10.840 min
Delta R.T.: -0.002 min
Response: 865982
Conc: 28.66 ng/ml




#7 Ethylbenzene

R.T.: 13.290 min
Delta R.T.: -0.002 min
Response: 195439
Conc: 7.64 ng/ml



#8 m-Xylene


R.T.: 13.421 min
Delta R.T.: -0.005 min
Response: 816548
Conc: 30.90 ng/ml



#9 O-Xylene

R.T.: 14.155 min
Delta R.T.: -0.002 min
Response: 326325
Conc: 12.05 ng/ml

Instrument :
FID_B
ClientSampleId :
RR-GAS-WP



#10 1,2,4-Trimethylbenzene

R.T.: 16.438 min
Delta R.T.: -0.002 min
Response: 364558
Conc: 13.95 ng/ml