

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB040519\
 Data File : FB020541.D
 Signal(s) : FID2B.CH
 Acq On : 5 Apr 2019 16:00
 Operator : MJ/SJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :
 20 PPB GRO STD

Integration File: Calibration.e
 Quant Time: Apr 06 01:17:35 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB040419.M
 Quant Title :
 QLast Update : Thu Apr 04 17:27:15 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.996	364543	19.764 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.866	1136126	32.273 ng/ml
2) t 2,2,4-Trimethylpentane	7.624	975735	32.728 ng/ml
3) t n-Heptane	7.953	422443	10.761 ng/ml
4) t Benzene	8.089	428722	10.438 ng/ml
6) t Toluene	10.840	955153	31.611 ng/ml
7) t Ethylbenzene	13.291	266717	10.426 ng/ml
8) t m-Xylene	13.424	551593	20.871 ng/ml
9) t O-Xylene	14.155	563601	20.819 ng/ml
10) t 1,2,4-Trimethylbenzene	16.439	563591	21.571 ng/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB040519\
Data File : FB020541.D
Signal(s) : FID2B.CH
Acq On : 5 Apr 2019 16:00
Operator : MJ/SJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :
20 PPB GRO STD

Integration File: Calibration.e
Quant Time: Apr 06 01:17:35 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB040419.M
Quant Title :
QLast Update : Thu Apr 04 17:27:15 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane


R.T.: 4.866 min
Delta R.T.: -0.003 min
Response: 1136126
Conc: 32.27 ng/ml

Instrument :
FID_B
ClientSampleId :
20 PPB GRO STD




#2 2,2,4-Trimethylpentane

R.T.: 7.624 min
Delta R.T.: -0.003 min
Response: 975735
Conc: 32.73 ng/ml




#3 n-Heptane

R.T.: 7.953 min
Delta R.T.: -0.002 min
Response: 422443
Conc: 10.76 ng/ml



#4 Benzene


R.T.: 8.089 min
Delta R.T.: -0.002 min
Response: 428722
Conc: 10.44 ng/ml



#5 AAA-TFT


R.T.: 8.996 min
Delta R.T.: -0.002 min
Response: 364543
Conc: 19.76 ng/ml

Instrument :
FID_B
ClientSampleId :
20 PPB GRO STD




#6 Toluene

R.T.: 10.840 min
Delta R.T.: -0.002 min
Response: 955153
Conc: 31.61 ng/ml




#7 Ethylbenzene

R.T.: 13.291 min
Delta R.T.: -0.002 min
Response: 266717
Conc: 10.43 ng/ml



#8 m-Xylene


R.T.: 13.424 min
Delta R.T.: -0.002 min
Response: 551593
Conc: 20.87 ng/ml



#9 O-Xylene

R.T.: 14.155 min
Delta R.T.: -0.002 min
Response: 563601
Conc: 20.82 ng/ml

Instrument :
FID_B
ClientSampleId :
20 PPB GRO STD



#10 1,2,4-Trimethylbenzene

R.T.: 16.439 min
Delta R.T.: -0.001 min
Response: 563591
Conc: 21.57 ng/ml