

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB040519\
 Data File : FB020574.D
 Signal(s) : FID2B.CH
 Acq On : 6 Apr 2019 9:02
 Operator : MJ/SJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :

Integration File: SAMPLE.e
 Quant Time: Apr 08 05:31:11 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB040419.M
 Quant Title :
 QLast Update : Thu Apr 04 17:27:15 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.995	352873	19.132 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.865	1047532	29.756 ng/ml
2) t 2,2,4-Trimethylpentane	7.622	896366	30.066 ng/ml
3) t n-Heptane	7.952	373520	9.515 ng/ml
4) t Benzene	8.087	423277	10.305 ng/ml
6) t Toluene	10.838	925803	30.639 ng/ml
7) t Ethylbenzene	13.289	258524	10.106 ng/ml
8) t m-Xylene	13.422	520193	19.682 ng/ml
9) t O-Xylene	14.153	543163	20.064 ng/ml
10) t 1,2,4-Trimethylbenzene	16.437	523241	20.027 ng/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB040519\
Data File : FB020574.D
Signal(s) : FID2B.CH
Acq On : 6 Apr 2019 9:02
Operator : MJ/SJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 43 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :

Integration File: SAMPLE.e
Quant Time: Apr 08 05:31:11 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB040419.M
Quant Title :
QLast Update : Thu Apr 04 17:27:15 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane


R.T.: 4.865 min
Delta R.T.: -0.004 min
Response: 1047532
Conc: 29.76 ng/ml

Instrument :
FID_B
ClientSampleId :




#2 2,2,4-Trimethylpentane

R.T.: 7.622 min
Delta R.T.: -0.005 min
Response: 896366
Conc: 30.07 ng/ml




#3 n-Heptane

R.T.: 7.952 min
Delta R.T.: -0.004 min
Response: 373520
Conc: 9.51 ng/ml



#4 Benzene


R.T.: 8.087 min
Delta R.T.: -0.004 min
Response: 423277
Conc: 10.31 ng/ml



#5 AAA-TFT


R.T.: 8.995 min
Delta R.T.: -0.003 min
Response: 352873
Conc: 19.13 ng/ml

Instrument :
FID_B
ClientSampleId :




#6 Toluene

R.T.: 10.838 min
Delta R.T.: -0.004 min
Response: 925803
Conc: 30.64 ng/ml




#7 Ethylbenzene

R.T.: 13.289 min
Delta R.T.: -0.004 min
Response: 258524
Conc: 10.11 ng/ml



#8 m-Xylene


R.T.: 13.422 min
Delta R.T.: -0.004 min
Response: 520193
Conc: 19.68 ng/ml



#9 O-Xylene

R.T.: 14.153 min
Delta R.T.: -0.004 min
Response: 543163
Conc: 20.06 ng/ml

Instrument :
FID_B
ClientSampleId :



#10 1,2,4-Trimethylbenzene

R.T.: 16.437 min
Delta R.T.: -0.004 min
Response: 523241
Conc: 20.03 ng/ml