

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB040820\
 Data File : FB024110.D
 Signal(s) : FID2B.CH
 Acq On : 8 Apr 2020 16:08
 Operator : DD/AJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :
 20 PPB GRO STD

Integration File: Calibration.e
 Quant Time: Apr 09 00:38:43 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB040820.M
 Quant Title :
 QLast Update : Wed Apr 08 10:11:26 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.801	492615	19.662 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.691	960726	29.432 ng/ml
2) t 2,2,4-Trimethylpentane	7.420	1010532	29.611 ng/ml
3) t n-Heptane	7.760	331275	9.815 ng/ml
4) t Benzene	7.894	400198	9.916 ng/ml
6) t Toluene	10.633	1158112	29.460 ng/ml
7) t Ethylbenzene	13.073	337275	9.252 ng/ml
8) t m-Xylene	13.207	665777	19.019 ng/ml
9) t O-Xylene	13.935	654836	19.213 ng/ml
10) t 1,2,4-Trimethylbenzene	16.212	555608	19.568 ng/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB040820\
Data File : FB024110.D
Signal(s) : FID2B.CH
Acq On : 8 Apr 2020 16:08
Operator : DD/AJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :
20 PPB GRO STD

Integration File: Calibration.e
Quant Time: Apr 09 00:38:43 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB040820.M
Quant Title :
QLast Update : Wed Apr 08 10:11:26 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane


R.T.: 4.691 min
Delta R.T.: 0.000 min
Response: 960726
Conc: 29.43 ng/ml

Instrument :
FID_B
ClientSampleId :
20 PPB GRO STD




#2 2,2,4-Trimethylpentane

R.T.: 7.420 min
Delta R.T.: 0.000 min
Response: 1010532
Conc: 29.61 ng/ml




#3 n-Heptane

R.T.: 7.760 min
Delta R.T.: 0.000 min
Response: 331275
Conc: 9.82 ng/ml



#4 Benzene


R.T.: 7.894 min
Delta R.T.: 0.000 min
Response: 400198
Conc: 9.92 ng/ml



#5 AAA-TFT


R.T.: 8.801 min
Delta R.T.: 0.000 min
Response: 492615
Conc: 19.66 ng/ml

Instrument :
FID_B
ClientSampleId :
20 PPB GRO STD




#6 Toluene

R.T.: 10.633 min
Delta R.T.: 0.001 min
Response: 1158112
Conc: 29.46 ng/ml




#7 Ethylbenzene

R.T.: 13.073 min
Delta R.T.: 0.001 min
Response: 337275
Conc: 9.25 ng/ml



#8 m-Xylene


R.T.: 13.207 min
Delta R.T.: 0.000 min
Response: 665777
Conc: 19.02 ng/ml



#9 O-Xylene

R.T.: 13.935 min
Delta R.T.: 0.001 min
Response: 654836
Conc: 19.21 ng/ml

Instrument :
FID_B
ClientSampleId :
20 PPB GRO STD



#10 1,2,4-Trimethylbenzene

R.T.: 16.212 min
Delta R.T.: 0.000 min
Response: 555608
Conc: 19.57 ng/ml