

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB041020\
 Data File : FB024111.D
 Signal(s) : FID2B.CH
 Acq On : 10 Apr 2020 6:53
 Operator : DD/AJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Apr 11 00:43:39 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB040820.M
 Quant Title :
 QLast Update : Wed Apr 08 10:11:26 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.797	426670	17.030 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.690	964535	29.549 ng/ml
2) t 2,2,4-Trimethylpentane	7.416	1004562	29.436 ng/ml
3) t n-Heptane	7.755	337784	10.008 ng/ml
4) t Benzene	7.890	419734	10.400 ng/ml
6) t Toluene	10.628	1174096	29.866 ng/ml
7) t Ethylbenzene	13.069	353728	9.703 ng/ml
8) t m-Xylene	13.202	687982	19.653 ng/ml
9) t O-Xylene	13.930	675022	19.805 ng/ml
10) t 1,2,4-Trimethylbenzene	16.209	569072	20.042 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB041020\
Data File : FB024111.D
Signal(s) : FID2B.CH
Acq On : 10 Apr 2020 6:53
Operator : DD/AJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 1 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Apr 11 00:43:39 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB040820.M
Quant Title :
QLast Update : Wed Apr 08 10:11:26 2020
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.690 min
Delta R.T.: -0.001 min
Response: 964535
Conc: 29.55 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.416 min
Delta R.T.: -0.004 min
Response: 1004562
Conc: 29.44 ng/ml




#3 n-Heptane

R.T.: 7.755 min
Delta R.T.: -0.004 min
Response: 337784
Conc: 10.01 ng/ml




#4 Benzene

R.T.: 7.890 min
Delta R.T.: -0.004 min
Response: 419734
Conc: 10.40 ng/ml




#5 AAA-TFT

R.T.: 8.797 min
Delta R.T.: -0.004 min
Response: 426670
Conc: 17.03 ng/ml




#6 Toluene

R.T.: 10.628 min
Delta R.T.: -0.004 min
Response: 1174096
Conc: 29.87 ng/ml




#7 Ethylbenzene

R.T.: 13.069 min
Delta R.T.: -0.003 min
Response: 353728
Conc: 9.70 ng/ml




#8 m-Xylene

R.T.: 13.202 min
Delta R.T.: -0.003 min
Response: 687982
Conc: 19.65 ng/ml



#9 O-Xylene

R.T.: 13.930 min
Delta R.T.: -0.003 min
Response: 675022
Conc: 19.81 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.209 min
Delta R.T.: -0.002 min
Response: 569072
Conc: 20.04 ng/ml