

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB041020\
 Data File : FB024126.D
 Signal(s) : FID2B.CH
 Acq On : 10 Apr 2020 14:29
 Operator : DD/AJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Apr 11 00:45:33 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB040820.M
 Quant Title :
 QLast Update : Wed Apr 08 10:11:26 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units


System Monitoring Compounds			
5) s AAA-TFT	8.799	528224	21.084 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.690	975486	29.884 ng/ml
2) t 2,2,4-Trimethylpentane	7.418	1027503	30.108 ng/ml
3) t n-Heptane	7.759	339777	10.067 ng/ml
4) t Benzene	7.893	421889	10.453 ng/ml
6) t Toluene	10.631	1201081	30.553 ng/ml
7) t Ethylbenzene	13.072	373724	10.252 ng/ml
8) t m-Xylene	13.205	726028	20.740 ng/ml
9) t O-Xylene	13.933	707947	20.771 ng/ml
10) t 1,2,4-Trimethylbenzene	16.211	581231	20.471 ng/ml


(f)=RT Delta > 1/2 Window

(m)=manual int.

Integration File: Calibration.e
Quant Time: Apr 11 00:45:33 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB040820.M
Quant Title :
QLast Update : Wed Apr 08 10:11:26 2020
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.690 min
Delta R.T.: 0.000 min
Response: 975486
Conc: 29.88 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.418 min
Delta R.T.: -0.002 min
Response: 1027503
Conc: 30.11 ng/ml




#3 n-Heptane

R.T.: 7.759 min
Delta R.T.: 0.000 min
Response: 339777
Conc: 10.07 ng/ml




#4 Benzene

R.T.: 7.893 min
Delta R.T.: 0.000 min
Response: 421889
Conc: 10.45 ng/ml




#5 AAA-TFT

R.T.: 8.799 min
Delta R.T.: -0.002 min
Response: 528224
Conc: 21.08 ng/ml




#6 Toluene

R.T.: 10.631 min
Delta R.T.: 0.000 min
Response: 1201081
Conc: 30.55 ng/ml




#7 Ethylbenzene

R.T.: 13.072 min
Delta R.T.: 0.000 min
Response: 373724
Conc: 10.25 ng/ml




#8 m-Xylene

R.T.: 13.205 min
Delta R.T.: 0.000 min
Response: 726028
Conc: 20.74 ng/ml



#9 O-Xylene

R.T.: 13.933 min
Delta R.T.: 0.000 min
Response: 707947
Conc: 20.77 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.211 min
Delta R.T.: 0.000 min
Response: 581231
Conc: 20.47 ng/ml