

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB041219\
Data File : FB020660.D
Signal(s) : FID2B.CH
Acq On : 12 Apr 2019 8:53
Operator : MJ/SJ
Sample : VBF0412S1
Misc : 5.02g/5.00ml DI WATER
ALS Vial : 2 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Apr 13 00:06:50 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB040419.M
Quant Title :
QLast Update : Thu Apr 04 17:27:15 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um

Compound		R.T.	Response	Conc Units

System Monitoring Compounds				
5) s	AAA-TFT	8.994	331962	17.998 ng/ml
Target Compounds				
2) t	2,2,4-Trimethylpentane	7.619	4294	0.144 ng/ml
6) t	Toluene	10.839	5251	0.174 ng/ml
9) t	O-Xylene	14.140	7752	0.286 ng/ml
10) t	1,2,4-Trimethylbenzene	16.444	7983	0.306 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB041219\
Data File : FB020660.D
Signal(s) : FID2B.CH
Acq On : 12 Apr 2019 8:53
Operator : MJ/SJ
Sample : VBF0412S1
Misc : 5.02g/5.00ml DI WATER
ALS Vial : 2 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Apr 13 00:06:50 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB040419.M
Quant Title :
QLast Update : Thu Apr 04 17:27:15 2019
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#2 2,2,4-Trimethylpentane

R.T.: 7.619 min
Delta R.T.: -0.008 min
Response: 4294
Conc: 0.14 ng/ml




#5 AAA-TFT

R.T.: 8.994 min
Delta R.T.: -0.004 min
Response: 331962
Conc: 18.00 ng/ml




#6 Toluene

R.T.: 10.839 min
Delta R.T.: -0.004 min
Response: 5251
Conc: 0.17 ng/ml



#9 O-Xylene

R.T.: 14.140 min
Delta R.T.: -0.016 min
Response: 7752
Conc: 0.29 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.444 min
Delta R.T.: 0.004 min
Response: 7983
Conc: 0.31 ng/ml