

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB041322\
Data File : FB027005.D
Signal(s) : FID2B.CH
Acq On : 13 Apr 2022 10:26
Operator : AJ/MA
Sample : VBF0413S2 50X
Misc : 5.01G/5.00 ML DI WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :
VBF0413S2

Integration File: Calibration.e
Quant Time: Apr 14 03:37:36 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB040922.M
Quant Title :
QLast Update : Sat Apr 09 04:06:25 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.791	233863	20.686 ng/ml
Target Compounds			

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB041322\
Data File : FB027005.D
Signal(s) : FID2B.CH
Acq On : 13 Apr 2022 10:26
Operator : AJ/MA
Sample : VBF0413S2 50X
Misc : 5.01G/5.00 ML DI WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :
VBF0413S2

Integration File: Calibration.e
Quant Time: Apr 14 03:37:36 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB040922.M
Quant Title :
QLast Update : Sat Apr 09 04:06:25 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um

