

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB041520\
 Data File : FB024146.D
 Signal(s) : FID2B.CH
 Acq On : 15 Apr 2020 21:58
 Operator : DD/AJ
 Sample : L2123-08MS
 Misc : 5.03g/5.00ml DI WATER
 ALS Vial : 10 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Apr 16 01:58:01 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB040820.M
 Quant Title :
 QLast Update : Wed Apr 08 10:11:26 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.800	490118	19.563 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.691	869439	26.635 ng/ml
2) t 2,2,4-Trimethylpentane	7.419	948780	27.802 ng/ml
3) t n-Heptane	7.759	310199	9.191 ng/ml
4) t Benzene	7.894	380266	9.422 ng/ml
6) t Toluene	10.631	1071941	27.268 ng/ml
7) t Ethylbenzene	13.072	316531	8.683 ng/ml
8) t m-Xylene	13.205	616241	17.604 ng/ml
9) t O-Xylene	13.932	604984	17.750 ng/ml
10) t 1,2,4-Trimethylbenzene	16.210	519464	18.295 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB041520\
Data File : FB024146.D
Signal(s) : FID2B.CH
Acq On : 15 Apr 2020 21:58
Operator : DD/AJ
Sample : L2123-08MS
Misc : 5.03g/5.00ml DI WATER
ALS Vial : 10 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Apr 16 01:58:01 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB040820.M
Quant Title :
QLast Update : Wed Apr 08 10:11:26 2020
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.691 min
Delta R.T.: 0.000 min
Response: 869439
Conc: 26.64 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.419 min
Delta R.T.: 0.000 min
Response: 948780
Conc: 27.80 ng/ml




#3 n-Heptane

R.T.: 7.759 min
Delta R.T.: 0.000 min
Response: 310199
Conc: 9.19 ng/ml




#4 Benzene

R.T.: 7.894 min
Delta R.T.: 0.000 min
Response: 380266
Conc: 9.42 ng/ml




#5 AAA-TFT

R.T.: 8.800 min
Delta R.T.: 0.000 min
Response: 490118
Conc: 19.56 ng/ml




#6 Toluene

R.T.: 10.631 min
Delta R.T.: 0.000 min
Response: 1071941
Conc: 27.27 ng/ml




#7 Ethylbenzene

R.T.: 13.072 min
Delta R.T.: 0.000 min
Response: 316531
Conc: 8.68 ng/ml




#8 m-Xylene

R.T.: 13.205 min
Delta R.T.: 0.000 min
Response: 616241
Conc: 17.60 ng/ml



#9 O-Xylene

R.T.: 13.932 min
Delta R.T.: 0.000 min
Response: 604984
Conc: 17.75 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.210 min
Delta R.T.: -0.001 min
Response: 519464
Conc: 18.30 ng/ml