

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB041520\
 Data File : FB024137.D
 Signal(s) : FID2B.CH
 Acq On : 15 Apr 2020 17:23
 Operator : DD/AJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Apr 16 01:56:34 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB040820.M
 Quant Title :
 QLast Update : Wed Apr 08 10:11:26 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.796	486599	19.422 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.688	899154	27.546 ng/ml
2) t 2,2,4-Trimethylpentane	7.414	944065	27.663 ng/ml
3) t n-Heptane	7.754	314281	9.312 ng/ml
4) t Benzene	7.889	400041	9.912 ng/ml
6) t Toluene	10.628	1151612	29.295 ng/ml
7) t Ethylbenzene	13.069	355201	9.743 ng/ml
8) t m-Xylene	13.202	683104	19.514 ng/ml
9) t O-Xylene	13.930	671743	19.709 ng/ml
10) t 1,2,4-Trimethylbenzene	16.208	566635	19.956 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB041520\
Data File : FB024137.D
Signal(s) : FID2B.CH
Acq On : 15 Apr 2020 17:23
Operator : DD/AJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 1 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Apr 16 01:56:34 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB040820.M
Quant Title :
QLast Update : Wed Apr 08 10:11:26 2020
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.688 min
Delta R.T.: -0.003 min
Response: 899154
Conc: 27.55 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.414 min
Delta R.T.: -0.006 min
Response: 944065
Conc: 27.66 ng/ml




#3 n-Heptane

R.T.: 7.754 min
Delta R.T.: -0.006 min
Response: 314281
Conc: 9.31 ng/ml




#4 Benzene

R.T.: 7.889 min
Delta R.T.: -0.005 min
Response: 400041
Conc: 9.91 ng/ml




#5 AAA-TFT

R.T.: 8.796 min
Delta R.T.: -0.005 min
Response: 486599
Conc: 19.42 ng/ml




#6 Toluene

R.T.: 10.628 min
Delta R.T.: -0.004 min
Response: 1151612
Conc: 29.29 ng/ml




#7 Ethylbenzene

R.T.: 13.069 min
Delta R.T.: -0.004 min
Response: 355201
Conc: 9.74 ng/ml




#8 m-Xylene

R.T.: 13.202 min
Delta R.T.: -0.004 min
Response: 683104
Conc: 19.51 ng/ml



#9 O-Xylene

R.T.: 13.930 min
Delta R.T.: -0.004 min
Response: 671743
Conc: 19.71 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.208 min
Delta R.T.: -0.003 min
Response: 566635
Conc: 19.96 ng/ml