

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB041621\
Data File : FB025566.D
Signal(s) : FID2B.CH
Acq On : 16 Apr 2021 16:09
Operator : DD/AJ
Sample : M1992-01
Misc : 5.02g/5.00ml DI WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :
AU-05-040921

Manual Integrations
APPROVED

Ankita
4/20/2021 9:15:12 AM

Integration File: Calibration.e
Quant Time: Apr 16 16:31:29 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB041621.M
Quant Title :
QLast Update : Fri Apr 16 14:20:46 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	9.401	377496	19.196 ng/mlm
Target Compounds			

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB041621\
Data File : FB025566.D
Signal(s) : FID2B.CH
Acq On : 16 Apr 2021 16:09
Operator : DD/AJ
Sample : M1992-01
Misc : 5.02g/5.00ml DI WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :
AU-05-040921

Manual Integrations
APPROVED

Ankita
4/20/2021 9:15:12 AM

Integration File: Calibration.e
Quant Time: Apr 16 16:31:29 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB041621.M
Quant Title :
QLast Update : Fri Apr 16 14:20:46 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um

