

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB041720\
Data File : FB024178.D
Signal(s) : FID2B.CH
Acq On : 17 Apr 2020 20:29
Operator : DD/AJ
Sample : L2327-23
Misc : 5.02g/5.00ml DI WATER
ALS Vial : 22 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Apr 18 01:07:46 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB040820.M
Quant Title :
QLast Update : Wed Apr 08 10:11:26 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.801	365365	14.583 ng/ml
Target Compounds			
4) t Benzene	7.893	182	0.004 ng/ml
6) t Toluene	10.629	2119	0.054 ng/ml
7) t Ethylbenzene	13.056	1358	0.037 ng/ml
8) t m-Xylene	13.191	245	0.007 ng/ml
9) t O-Xylene	13.921	442	0.013 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB041720\
Data File : FB024178.D
Signal(s) : FID2B.CH
Acq On : 17 Apr 2020 20:29
Operator : DD/AJ
Sample : L2327-23
Misc : 5.02g/5.00ml DI WATER
ALS Vial : 22 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Apr 18 01:07:46 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB040820.M
Quant Title :
QLast Update : Wed Apr 08 10:11:26 2020
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#4 Benzene

R.T.: 7.893 min
Delta R.T.: 0.000 min
Response: 182
Conc: 0.00 ng/ml




#5 AAA-TFT

R.T.: 8.801 min
Delta R.T.: 0.000 min
Response: 365365
Conc: 14.58 ng/ml




#6 Toluene

R.T.: 10.629 min
Delta R.T.: -0.003 min
Response: 2119
Conc: 0.05 ng/ml




#7 Ethylbenzene

R.T.: 13.056 min
Delta R.T.: -0.016 min
Response: 1358
Conc: 0.04 ng/ml



#8 m-Xylene

R.T.: 13.191 min
Delta R.T.: -0.015 min
Response: 245
Conc: 0.01 ng/ml



#9 O-Xylene

R.T.: 13.921 min
Delta R.T.: -0.012 min
Response: 442
Conc: 0.01 ng/ml