

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB041820\
 Data File : FB024192.D
 Signal(s) : FID2B.CH
 Acq On : 18 Apr 2020 15:09
 Operator : DD/AJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Apr 20 02:41:21 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB040820.M
 Quant Title :
 QLast Update : Wed Apr 08 10:11:26 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.801	567526	22.653 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.692	976988	29.930 ng/ml
2) t 2,2,4-Trimethylpentane	7.421	1074209	31.477 ng/ml
3) t n-Heptane	7.759	346092	10.254 ng/ml
4) t Benzene	7.894	420433	10.417 ng/ml
6) t Toluene	10.632	1213928	30.880 ng/ml
7) t Ethylbenzene	13.073	355999	9.765 ng/ml
8) t m-Xylene	13.205	699862	19.993 ng/ml
9) t O-Xylene	13.933	688094	20.189 ng/ml
10) t 1,2,4-Trimethylbenzene	16.211	576679	20.310 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB041820\
Data File : FB024192.D
Signal(s) : FID2B.CH
Acq On : 18 Apr 2020 15:09
Operator : DD/AJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 13 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Apr 20 02:41:21 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB040820.M
Quant Title :
QLast Update : Wed Apr 08 10:11:26 2020
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.692 min
Delta R.T.: 0.001 min
Response: 976988
Conc: 29.93 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.421 min
Delta R.T.: 0.001 min
Response: 1074209
Conc: 31.48 ng/ml




#3 n-Heptane

R.T.: 7.759 min
Delta R.T.: 0.000 min
Response: 346092
Conc: 10.25 ng/ml




#4 Benzene

R.T.: 7.894 min
Delta R.T.: 0.000 min
Response: 420433
Conc: 10.42 ng/ml




#5 AAA-TFT

R.T.: 8.801 min
Delta R.T.: 0.000 min
Response: 567526
Conc: 22.65 ng/ml




#6 Toluene

R.T.: 10.632 min
Delta R.T.: 0.000 min
Response: 1213928
Conc: 30.88 ng/ml




#7 Ethylbenzene

R.T.: 13.073 min
Delta R.T.: 0.000 min
Response: 355999
Conc: 9.77 ng/ml




#8 m-Xylene

R.T.: 13.205 min
Delta R.T.: 0.000 min
Response: 699862
Conc: 19.99 ng/ml



#9 O-Xylene

R.T.: 13.933 min
Delta R.T.: 0.000 min
Response: 688094
Conc: 20.19 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.211 min
Delta R.T.: 0.000 min
Response: 576679
Conc: 20.31 ng/ml