

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB041923\
 Data File : FB029712.D
 Signal(s) : FID2B.CH
 Acq On : 19 Apr 2023 21:44
 Operator : YP/AJ
 Sample : BSF0619S3
 Misc : 5.00G/5.00 ML DI WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :

Integration File: Calibration.e
 Quant Time: Apr 20 01:22:04 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB041723.M
 Quant Title :
 QLast Update : Mon Apr 17 12:32:26 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.783	234905	17.627 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.700	389160	27.660 ng/ml
2) t 2,2,4-Trimethylpentane	7.408	442454	28.804 ng/ml
3) t n-Heptane	7.743	148587	9.232 ng/ml
4) t Benzene	7.879	195420	9.841 ng/ml
6) t Toluene	10.612	573975	29.717 ng/ml
7) t Ethylbenzene	13.048	181280	10.016 ng/ml
8) t m-Xylene	13.181	359309	19.867 ng/ml
9) t O-Xylene	13.908	350292	19.976 ng/ml
10) t 1,2,4-Trimethylbenzene	16.183	346840	19.408 ng/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB041923\
Data File : FB029712.D
Signal(s) : FID2B.CH
Acq On : 19 Apr 2023 21:44
Operator : YP/AJ
Sample : BSF0619S3
Misc : 5.00G/5.00 ML DI WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :

Integration File: Calibration.e
Quant Time: Apr 20 01:22:04 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB041723.M
Quant Title :
QLast Update : Mon Apr 17 12:32:26 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane


R.T.: 4.700 min
Delta R.T.: -0.005 min
Response: 389160
Conc: 27.66 ng/ml

Instrument :
FID_B
ClientSampleId :




#2 2,2,4-Trimethylpentane

R.T.: 7.408 min
Delta R.T.: -0.005 min
Response: 442454
Conc: 28.80 ng/ml




#3 n-Heptane

R.T.: 7.743 min
Delta R.T.: -0.004 min
Response: 148587
Conc: 9.23 ng/ml



#4 Benzene


R.T.: 7.879 min
Delta R.T.: -0.004 min
Response: 195420
Conc: 9.84 ng/ml



#5 AAA-TFT


R.T.: 8.783 min
Delta R.T.: -0.004 min
Response: 234905
Conc: 17.63 ng/ml

Instrument :
FID_B
ClientSampleId :




#6 Toluene

R.T.: 10.612 min
Delta R.T.: -0.004 min
Response: 573975
Conc: 29.72 ng/ml




#7 Ethylbenzene

R.T.: 13.048 min
Delta R.T.: -0.004 min
Response: 181280
Conc: 10.02 ng/ml



#8 m-Xylene

R.T.: 13.181 min
Delta R.T.: -0.004 min
Response: 359309
Conc: 19.87 ng/ml




#9 O-Xylene

R.T.: 13.908 min
Delta R.T.: -0.004 min
Response: 350292
Conc: 19.98 ng/ml

Instrument :

FID_B

ClientSampleId :



#10 1,2,4-Trimethylbenzene

R.T.: 16.183 min
Delta R.T.: -0.003 min
Response: 346840
Conc: 19.41 ng/ml