

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB042120\
 Data File : FB024230.D
 Signal(s) : FID2B.CH
 Acq On : 22 Apr 2020 1:34
 Operator : DD/AJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Apr 22 03:07:32 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB040820.M
 Quant Title :
 QLast Update : Wed Apr 08 10:11:26 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.798	478286	19.091 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.691	869742	26.645 ng/ml
2) t 2,2,4-Trimethylpentane	7.417	966455	28.320 ng/ml
3) t n-Heptane	7.757	307978	9.125 ng/ml
4) t Benzene	7.892	383784	9.509 ng/ml
6) t Toluene	10.630	1099923	27.980 ng/ml
7) t Ethylbenzene	13.070	340260	9.334 ng/ml
8) t m-Xylene	13.203	661308	18.891 ng/ml
9) t O-Xylene	13.931	648082	19.015 ng/ml
10) t 1,2,4-Trimethylbenzene	16.209	523414	18.434 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB042120\
Data File : FB024230.D
Signal(s) : FID2B.CH
Acq On : 22 Apr 2020 1:34
Operator : DD/AJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 16 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Apr 22 03:07:32 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB040820.M
Quant Title :
QLast Update : Wed Apr 08 10:11:26 2020
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.691 min
Delta R.T.: 0.000 min
Response: 869742
Conc: 26.64 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.417 min
Delta R.T.: -0.003 min
Response: 966455
Conc: 28.32 ng/ml




#3 n-Heptane

R.T.: 7.757 min
Delta R.T.: -0.003 min
Response: 307978
Conc: 9.12 ng/ml




#4 Benzene

R.T.: 7.892 min
Delta R.T.: -0.002 min
Response: 383784
Conc: 9.51 ng/ml




#5 AAA-TFT

R.T.: 8.798 min
Delta R.T.: -0.003 min
Response: 478286
Conc: 19.09 ng/ml




#6 Toluene

R.T.: 10.630 min
Delta R.T.: -0.002 min
Response: 1099923
Conc: 27.98 ng/ml




#7 Ethylbenzene

R.T.: 13.070 min
Delta R.T.: -0.002 min
Response: 340260
Conc: 9.33 ng/ml




#8 m-Xylene

R.T.: 13.203 min
Delta R.T.: -0.002 min
Response: 661308
Conc: 18.89 ng/ml



#9 O-Xylene

R.T.: 13.931 min
Delta R.T.: -0.002 min
Response: 648082
Conc: 19.01 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.209 min
Delta R.T.: -0.002 min
Response: 523414
Conc: 18.43 ng/ml